

## LABORATORY REPORT

If you have any questions concerning this report, please do not hesitate to call us at (800) 332-4345 or (574) 233-4777.

*This report may not be reproduced, except in full, without written approval from EEA.*

## STATE CERTIFICATION LIST

State	Certification	State	Certification
Alabama	40700	Montana	CERT0026
Alaska	IN00035	Nebraska	NE-OS-05-04
Arizona	AZ0432	Nevada	IN00035
Arkansas	IN00035	<a href="#">New Hampshire*</a>	<a href="#">2124</a>
California	2920	<a href="#">New Jersey*</a>	<a href="#">IN598</a>
Colorado	IN035	New Mexico	IN00035
Colorado Radiochemistry	IN035	<a href="#">New York*</a>	<a href="#">11398</a>
Connecticut	PH-0132	North Carolina	18700
Delaware	IN035	North Dakota	R-035
<a href="#">Florida*</a>	<a href="#">E87775</a>	Ohio	87775
Georgia	929	Oklahoma	D9508
Hawaii	IN035	<a href="#">Oregon (Primary AB)*</a>	<a href="#">4074-001</a>
Idaho	IN00035	<a href="#">Pennsylvania*</a>	<a href="#">68-00466</a>
<a href="#">Illinois*</a>	<a href="#">200001</a>	Puerto Rico	IN00035
Illinois Microbiology	17767	Rhode Island	LAO00343
Indiana Chemistry	C-71-01	South Carolina	95005
Indiana Microbiology	M-76-07	South Dakota	IN00035
Iowa	098	Tennessee	TN02973
<a href="#">Kansas*</a>	<a href="#">E-10233</a>	<a href="#">Texas*</a>	T104704187-15-8
Kentucky	90056	Texas/TCEQ	TX207
<a href="#">Louisiana*</a>	<a href="#">LA170006</a>	<a href="#">Utah*</a>	<a href="#">IN00035</a>
Maine	IN00035	Vermont	VT-8775
Maryland	209	<a href="#">Virginia*</a>	<a href="#">460275</a>
Massachusetts	M-IN035	Washington	C837
Michigan	9926	West Virginia	9927 C
<a href="#">Minnesota*</a>	<a href="#">018-999-338</a>	Wisconsin	999766900
Mississippi	IN035	Wyoming	IN035
Missouri	880		

[\\*NELAP/TNI Recognized Accreditation Bodies](#)

110 South Hill Street  
 South Bend, IN 46617  
 Tel: (574) 233-4777  
 Fax: (574) 233-8207  
 1 800 332 4345

## Laboratory Report

Client: C.T. Male Associates  
 Attn: Kirk Moline  
 50 Century Hill Drive  
 Latham, NY 12110

Report: 392201  
 Priority: Standard Written  
 Status: Final

Project: 14.4756

### SUMMARY OF DETECTIONS

Parameter	Method	Result	Units	Run #
Perfluorobutanesulfonic acid (PFBS)	537	2.1	ng/L	231772
Perfluoroheptanoic acid (PFHpA)	537	17	ng/L	231772
Perfluorohexanoic acid (PFHxA)	537	14	ng/L	231772
Perfluorooctane sulfonate (PFOS)	537	2.9	ng/L	231772
Perfluorooctanoic acid (PFOA)	537	700	ng/L	231772

Note: The results presented relate only to the samples provided for analysis.

We appreciate the opportunity to provide you with this analysis. If you have any questions concerning this report, please do not hesitate to call Kelly Trott at (574) 233-4777.

Note: This report may not be reproduced, except in full, without written approval from EEA.

*Dana M. Siz* Reporter

Reviewed By

*Kelly Trott* Analytical Services Manager

Title

08/03/2017

Date

08/03/2017

Finalized By

Title

Date

110 South Hill Street  
 South Bend, IN 46617  
 Tel: (574) 233-4777  
 Fax: (574) 233-8207  
 1 800 332 4345

## Laboratory Report

Client: C.T. Male Associates  
  
 Attn: Kirk Moline  
 50 Century Hill Drive  
 Latham, NY 12110

Report: 392201  
 Priority: Standard Written  
 Status: Final  
 PWS ID: Not Supplied

Sample Information					
EEA ID #	Client ID	Method	Collected Date / Time	Collected By:	Received Date / Time
3726944	GAC Influent	537	07/06/17 09:30	Client	07/07/17 08:30
3726945	GAC Midfluent	537	07/06/17 09:33	Client	07/07/17 08:30
3726946	GAC Effluent	537	07/06/17 09:35	Client	07/07/17 08:30
3726947	FTB	537	07/06/17 09:37	Client	07/07/17 08:30
3726948	LTB 6/13/17	537	07/06/17 09:30	Client	07/07/17 08:30

### Report Summary

Project: 14.4756

Detailed quantitative results are presented on the following pages. The results presented relate only to the samples provided for analysis.

We appreciate the opportunity to provide you with this analysis. If you have any questions concerning this report, please do not hesitate to call Kelly Trott at (574) 233-4777.

*Note: This report may not be reproduced, except in full, without written approval from EEA.*

*Kelly Trott Analytical Services Manager*

Authorized Signature

Title

08/03/2017

Date

Client Name: C.T. Male Associates

Report #: 392201

Sampling Point: GAC Influent

PWS ID: Not Supplied

EEA Methods									
Analyte ID #	Analyte	Method	Reg Limit	MRL†	Result	Units	Preparation Date	Analyzed Date	EEA ID #
375-73-5	Perfluorobutanesulfonic acid (PFBS)	537	---	2.0	2.1	ng/L	07/11/17 06:59	07/11/17 23:42	3726944
335-76-2	Perfluorodecanoic acid (PFDA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/11/17 23:42	3726944
375-85-9	Perfluoroheptanoic acid (PFHpA)	537	---	2.0	17	ng/L	07/11/17 06:59	07/11/17 23:42	3726944
355-46-4	Perfluorohexanesulfonic acid (PFHxS)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/11/17 23:42	3726944
307-24-4	Perfluorohexanoic acid (PFHxA)	537	---	2.0	14	ng/L	07/11/17 06:59	07/11/17 23:42	3726944
307-55-1	Perfluorolauric acid (PFDoA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/11/17 23:42	3726944
376-06-7	Perfluoromyristic acid (PFTA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/11/17 23:42	3726944
375-95-1	Perfluorononanoic acid (PFNA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/11/17 23:42	3726944
1763-23-1	Perfluorooctane sulfonate (PFOS)	537	---	2.0	2.9	ng/L	07/11/17 06:59	07/11/17 23:42	3726944
335-67-1	Perfluorooctanoic acid (PFOA)	537	---	2.0	700	ng/L	07/11/17 06:59	07/12/17 09:25	3726944
72629-94-8	Perfluorotridecanoic acid (PFTTrDA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/11/17 23:42	3726944
2058-94-8	Perfluoroundecanoic acid (PFUnA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/11/17 23:42	3726944

Sampling Point: GAC Midfluent

PWS ID: Not Supplied

EEA Methods									
Analyte ID #	Analyte	Method	Reg Limit	MRL†	Result	Units	Preparation Date	Analyzed Date	EEA ID #
375-73-5	Perfluorobutanesulfonic acid (PFBS)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/11/17 23:59	3726945
335-76-2	Perfluorodecanoic acid (PFDA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/11/17 23:59	3726945
375-85-9	Perfluoroheptanoic acid (PFHpA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/11/17 23:59	3726945
355-46-4	Perfluorohexanesulfonic acid (PFHxS)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/11/17 23:59	3726945
307-24-4	Perfluorohexanoic acid (PFHxA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/11/17 23:59	3726945
307-55-1	Perfluorolauric acid (PFDoA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/11/17 23:59	3726945
376-06-7	Perfluoromyristic acid (PFTA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/11/17 23:59	3726945
375-95-1	Perfluorononanoic acid (PFNA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/11/17 23:59	3726945
1763-23-1	Perfluorooctane sulfonate (PFOS)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/11/17 23:59	3726945
335-67-1	Perfluorooctanoic acid (PFOA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/11/17 23:59	3726945
72629-94-8	Perfluorotridecanoic acid (PFTTrDA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/11/17 23:59	3726945
2058-94-8	Perfluoroundecanoic acid (PFUnA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/11/17 23:59	3726945

Sampling Point: GAC Effluent

PWS ID: Not Supplied

EEA Methods									
Analyte ID #	Analyte	Method	Reg Limit	MRL†	Result	Units	Preparation Date	Analyzed Date	EEA ID #
375-73-5	Perfluorobutanesulfonic acid (PFBS)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:33	3726946
335-76-2	Perfluorodecanoic acid (PFDA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:33	3726946
375-85-9	Perfluoroheptanoic acid (PFHpA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:33	3726946
355-46-4	Perfluorohexanesulfonic acid (PFHxS)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:33	3726946
307-24-4	Perfluorohexanoic acid (PFHxA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:33	3726946
307-55-1	Perfluorolauric acid (PFDoA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:33	3726946
376-06-7	Perfluoromyristic acid (PFTA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:33	3726946
375-95-1	Perfluorononanoic acid (PFNA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:33	3726946
1763-23-1	Perfluorooctane sulfonate (PFOS)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:33	3726946
335-67-1	Perfluorooctanoic acid (PFOA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:33	3726946
72629-94-8	Perfluorotridecanoic acid (PFTTrDA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:33	3726946
2058-94-8	Perfluoroundecanoic acid (PFUnA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:33	3726946

Sampling Point: FTB

PWS ID: Not Supplied

EEA Methods									
Analyte ID #	Analyte	Method	Reg Limit	MRL†	Result	Units	Preparation Date	Analyzed Date	EEA ID #
375-73-5	Perfluorobutanesulfonic acid (PFBS)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:49	3726947
335-76-2	Perfluorodecanoic acid (PFDA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:49	3726947
375-85-9	Perfluoroheptanoic acid (PFHpA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:49	3726947
355-46-4	Perfluorohexanesulfonic acid (PFHxS)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:49	3726947
307-24-4	Perfluorohexanoic acid (PFHxA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:49	3726947
307-55-1	Perfluorolauric acid (PFDoA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:49	3726947
376-06-7	Perfluoromyristic acid (PFTA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:49	3726947
375-95-1	Perfluorononanoic acid (PFNA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:49	3726947
1763-23-1	Perfluorooctane sulfonate (PFOS)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:49	3726947
335-67-1	Perfluorooctanoic acid (PFOA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:49	3726947
72629-94-8	Perfluorotridecanoic acid (PFTTrDA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:49	3726947
2058-94-8	Perfluoroundecanoic acid (PFUnA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 00:49	3726947

Sampling Point: LTB 6/13/17

PWS ID: Not Supplied

EEA Methods									
Analyte ID #	Analyte	Method	Reg Limit	MRL†	Result	Units	Preparation Date	Analyzed Date	EEA ID #
375-73-5	Perfluorobutanesulfonic acid (PFBS)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 01:06	3726948
335-76-2	Perfluorodecanoic acid (PFDA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 01:06	3726948
375-85-9	Perfluoroheptanoic acid (PFHpA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 01:06	3726948
355-46-4	Perfluorohexanesulfonic acid (PFHxS)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 01:06	3726948
307-24-4	Perfluorohexanoic acid (PFHxA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 01:06	3726948
307-55-1	Perfluorolauric acid (PFDoA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 01:06	3726948
376-06-7	Perfluoromyristic acid (PFTA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 01:06	3726948
375-95-1	Perfluorononanoic acid (PFNA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 01:06	3726948
1763-23-1	Perfluorooctane sulfonate (PFOS)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 01:06	3726948
335-67-1	Perfluorooctanoic acid (PFOA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 01:06	3726948
72629-94-8	Perfluorotridecanoic acid (PFTrDA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 01:06	3726948
2058-94-8	Perfluoroundecanoic acid (PFUnA)	537	---	2.0	< 2.0	ng/L	07/11/17 06:59	07/12/17 01:06	3726948

† EEA has demonstrated it can achieve these report limits in reagent water, but can not document them in all sample matrices.

Reg Limit Type:	MCL	SMCL	AL
Symbol:	*	^	!

## Lab Definitions

**Continuing Calibration Check Standard (CCC) / Continuing Calibration Verification (CCV) / Initial Calibration Verification Standard (ICV) / Initial Performance Check (IPC)** - is a standard containing one or more of the target analytes that is prepared from the same standards used to calibrate the instrument. This standard is used to verify the calibration curve at the beginning of each analytical sequence, and may also be analyzed throughout and at the end of the sequence. The concentration of continuing standards may be varied, when prescribed by the reference method, so that the range of the calibration curve is verified on a regular basis. CCL, CCM, and CCH are the CCC standards at low, mid, and high concentration levels, respectively.

**Internal Standards (IS)** - are pure compounds with properties similar to the analytes of interest, which are added to field samples or extracts, calibration standards, and quality control standards at a known concentration. They are used to measure the relative responses of the analytes of interest and surrogates in the sample, calibration standard or quality control standard.

**Laboratory Duplicate (LD)** - is a field sample aliquot taken from the same sample container in the laboratory and analyzed separately using identical procedures. Analysis of laboratory duplicates provides a measure of the precision of the laboratory procedures.

**Laboratory Fortified Blank (LFB) / Laboratory Control Sample (LCS)** - is an aliquot of reagent water to which known concentrations of the analytes of interest are added. The LFB is analyzed exactly the same as the field samples. LFBs are used to determine whether the method is in control. FBL, FBM, and FBH are the LFB samples at low, mid, and high concentration levels, respectively.

**Laboratory Method Blank (LMB) / Laboratory Reagent Blank (LRB)** - is a sample of reagent water included in the sample batch analyzed in the same way as the associated field samples. The LMB is used to determine if method analytes or other background contamination have been introduced during the preparation or analytical procedure. The LMB is analyzed exactly the same as the field samples.

**Laboratory Trip Blank (LTB) / Field Reagent Blank (FRB)** - is a sample of laboratory reagent water placed in a sample container in the laboratory and treated as a field sample, including storage, preservation, and all analytical procedures. The FRB/LTB container follows the collection bottles to and from the collection site, but the FRB/LTB is not opened at any time during the trip. The FRB/LTB is primarily a travel blank used to verify that the samples were not contaminated during shipment.

**Matrix Spike Duplicate Sample (MSD) / Laboratory Fortified Sample Matrix Duplicate (LFSMD)** - is a sample aliquot taken from the same field sample source as the Matrix Spike Sample to which known quantities of the analytes of interest are added in the laboratory. The MSD is analyzed exactly the same as the field samples. Analysis of the MSD provides a measure of the precision of the laboratory procedures in a specific matrix. SDL, SDM, and SDH / LFSMDL, LFSMDM, and LFSMDH are the MSD or LFSMD at low, mid, and high concentration levels, respectively.

**Matrix Spike Sample (MS) / Laboratory Fortified Sample Matrix (LFSM)** - is a sample aliquot taken from field sample source to which known quantities of the analytes of interest are added in the laboratory. The MS is analyzed exactly the same as the field samples. The purpose is to demonstrate recovery of the analytes from a sample matrix to determine if the specific matrix contributes bias to the analytical results. MSL, MSM, and MSH / LFSML, LFSMM, and LFSMH are the MS or LFSM at low, mid, and high concentration levels, respectively.

**Quality Control Standard (QCS) / Second Source Calibration Verification (SSCV)** - is a solution containing known concentrations of the analytes of interest prepared from a source different from the source of the calibration standards. The solution is obtained from a second manufacturer or lot if the lot can be demonstrated by the manufacturer as prepared independently from other lots. The QCS sample is analyzed using the same procedures as field samples. The QCS is used as a check on the calibration standards used in the method on a routine basis.

**Reporting Limit Check (RLC) / Initial Calibration Check Standard (ICCS)** - is a procedural standard that is analyzed each day to evaluate instrument performance at or below the minimum reporting limit (MRL).

**Surrogate Standard (SS) / Surrogate Analyte (SUR)** - is a pure compound with properties similar to the analytes of interest, which is highly unlikely to be found in any field sample, that is added to the field samples, calibration standards, blanks and quality control standards before sample preparation. The SS is used to evaluate the efficiency of the sample preparation process.