

ANALYTICAL REPORT

PREPARED FOR

Attn: Mr. Kirk Moline
CT Male Associates DPC
50 Century Hill Dr
Latham, New York 12110

Generated 4/25/2023 5:14:01 PM

JOB DESCRIPTION

Hoosick Falls WTP
SDG NUMBER HOO

JOB NUMBER

410-119243-1

Eurofins Lancaster Laboratories Environment Testing, LLC

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

Analytical test results meet all requirements of the associated regulatory program (i.e., NELAC (TNI), DoD, and ISO 17025) unless otherwise noted under the individual analysis.

Authorization



Generated
4/25/2023 5:14:01 PM

Authorized for release by
Paul Hobart, Project Manager
Paul.Hobart@et.eurofinsus.com
(617)312-8660

Compliance Statement

Analytical test results meet all requirements of the associated regulatory program (e.g., NELAC (TNI), DoD, and ISO 17025) unless otherwise noted under the individual analysis. Data qualifiers are applied to note exceptions. Noncompliant quality control (QC) is further explained in narrative comments.

- QC results that exceed the upper limits and are associated with non-detect samples are qualified but further narration is not required since the bias is high and does not change a non-detect result. Further narration is also not required with QC blank detection when the associated sample concentration is non-detect or more than ten times the level in the blank.
- Matrix QC may not be reported if insufficient sample or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD is performed, unless otherwise specified in the method.
- Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.

Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

Measurement uncertainty values, as applicable, are available upon request.

Test results relate only to the sample tested. Clients should be aware that a critical step in a chemical or microbiological analysis is the collection of the sample. Unless the sample analyzed is truly representative of the bulk of material involved, the test results will be meaningless. If you have questions regarding the proper techniques of collecting samples, please contact us. We cannot be held responsible for sample integrity, however, unless sampling has been performed by a member of our staff. Times are local to the area of activity. Parameters listed in the 40 CFR Part 136 Table II as "analyze immediately" and tested in the laboratory are not performed within 15 minutes of collection.

This report shall not be reproduced except in full, without the written approval of the laboratory.

WARRANTY AND LIMITS OF LIABILITY - In accepting analytical work, we warrant the accuracy of test results for the sample as submitted. The foregoing express warranty is exclusive and is given in lieu of all other warranties, expressed or implied, except as otherwise agreed. We disclaim any other warranties, expressed or implied, including a warranty of fitness for particular purpose and warranty of merchantability. In no event shall Eurofins Lancaster Laboratories Environmental, LLC be liable for indirect, special, consequential, or incidental damages including, but not limited to, damages for loss of profit or goodwill regardless of (A) the negligence (either sole or concurrent) of Eurofins Lancaster Laboratories Environmental and (B) whether Eurofins Lancaster Laboratories Environmental has been informed of the possibility of such damages. We accept no legal responsibility for the purposes for which the client uses the test results. Except as otherwise agreed, no purchase order or other order for work shall be accepted by Eurofins Lancaster Laboratories Environmental which includes any conditions that vary from the Standard Terms and Conditions, and Eurofins Lancaster Laboratories Environmental hereby objects to any conflicting terms contained in any acceptance or order submitted by client.




Table of Contents

Cover Page	1	3
Table of Contents	4	4
Definitions/Glossary	5	5
Case Narrative	6	6
Detection Summary	7	6
Client Sample Results	8	7
Surrogate Summary	16	8
Isotope Dilution Summary	17	9
QC Sample Results	18	9
QC Association Summary	22	10
Lab Chronicle	24	11
Certification Summary	26	12
Method Summary	27	12
Sample Summary	28	13
Chain of Custody	29	13
Receipt Checklists	30	14
		15
		16

Definitions/Glossary

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1

SDG: HOO

Qualifiers

LCMS

Qualifier	Qualifier Description
cn	Refer to Case Narrative for further detail
S1-	Surrogate recovery exceeds control limits, low biased.
U	Indicates the analyte was analyzed for but not detected.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
□	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1
SDG: HOO

Job ID: 410-119243-1

Laboratory: Eurofins Lancaster Laboratories Environment Testing, LLC

Narrative

Job Narrative 410-119243-1

Receipt

The samples were received on 3/17/2023 10:24 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 3.9°C

PFAS

Method 537_DW: The recovery for the labeled isotope(s) in the following sample: FTB01_230316 (410-119243-7) and LTB01_230316 (410-119243-8) is outside the QC acceptance limits. The following action was taken: This sample was re-extracted outside of the required holding time and the recovery for labeled isotope(s) was within QC acceptance limits.

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Detection Summary

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1
SDG: HOO

Client Sample ID: GAC Influent

Lab Sample ID: 410-119243-1

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorobutanoic acid	2.8		1.8	ng/L	1	537 (Mod)	Total/NA	
Perfluoropentanoic acid	2.2		1.8	ng/L	1	537 (Mod)	Total/NA	
Perfluorohexanoic acid	9.8		1.8	ng/L	1	537 DW	Total/NA	
Perfluoroheptanoic acid	11		1.8	ng/L	1	537 DW	Total/NA	
Perfluoroctanesulfonic acid	3.4		1.8	ng/L	1	537 DW	Total/NA	
Perfluoroctanoic acid - DL	350		18	ng/L	10	537 DW	Total/NA	

Client Sample ID: GAC Midfluent

Lab Sample ID: 410-119243-2

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorobutanoic acid	4.6		1.7	ng/L	1	537 (Mod)	Total/NA	

Client Sample ID: GAC Effluent

Lab Sample ID: 410-119243-3

No Detections.

Client Sample ID: PV-1_25

Lab Sample ID: 410-119243-4

No Detections.

Client Sample ID: PV-1_50

Lab Sample ID: 410-119243-5

No Detections.

Client Sample ID: PV-1_75

Lab Sample ID: 410-119243-6

No Detections.

Client Sample ID: FTB01_230316

Lab Sample ID: 410-119243-7

No Detections.

Client Sample ID: LTB01_230316

Lab Sample ID: 410-119243-8

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Lancaster Laboratories Environment Testing, LLC

Client Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1
SDG: HOO

Client Sample ID: GAC Influent

Lab Sample ID: 410-119243-1

Matrix: Water

Date Collected: 03/16/23 09:30
Date Received: 03/17/23 10:24

Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L	04/07/23 09:19	04/14/23 18:37		1
8:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L	04/07/23 09:19	04/14/23 18:37		1
Perfluorobutanoic acid	2.8		1.8	ng/L	04/07/23 09:19	04/14/23 18:37		1
Perfluorodecanesulfonic acid	1.8	U	1.8	ng/L	04/07/23 09:19	04/14/23 18:37		1
Perfluoroheptanesulfonic acid	1.8	U	1.8	ng/L	04/07/23 09:19	04/14/23 18:37		1
Perfluorooctanesulfonamide	1.8	U	1.8	ng/L	04/07/23 09:19	04/14/23 18:37		1
Perfluoropentanoic acid	2.2		1.8	ng/L	04/07/23 09:19	04/14/23 18:37		1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
M2-6:2 FTS	106		17 - 200			04/07/23 09:19	04/14/23 18:37	1
M2-8:2 FTS	134		33 - 200			04/07/23 09:19	04/14/23 18:37	1
13C4 PFBA	118		42 - 165			04/07/23 09:19	04/14/23 18:37	1
13C5 PFPeA	123		38 - 187			04/07/23 09:19	04/14/23 18:37	1
13C8 PFOS	115		51 - 159			04/07/23 09:19	04/14/23 18:37	1
13C8 FOSA	92		10 - 168			04/07/23 09:19	04/14/23 18:37	1
13C3 PFHxA	135		28 - 188			04/07/23 09:19	04/14/23 18:37	1

Method: EPA 537 DW - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	9.8		1.8	ng/L	03/20/23 11:02	04/11/23 18:21		1
Perfluoroheptanoic acid	11		1.8	ng/L	03/20/23 11:02	04/11/23 18:21		1
Perfluorononanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 18:21		1
Perfluorodecanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 18:21		1
Perfluorotridecanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 18:21		1
Perfluorotetradecanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 18:21		1
Perfluorobutanesulfonic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 18:21		1
Perfluorohexanesulfonic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 18:21		1
Perfluorooctanesulfonic acid	3.4		1.8	ng/L	03/20/23 11:02	04/11/23 18:21		1
NEtFOSAA	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 18:21		1
NMeFOSAA	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 18:21		1
Perfluoroundecanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 18:21		1
Perfluorododecanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 18:21		1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	91		70 - 130			03/20/23 11:02	04/11/23 18:21	1
13C2 PFDA	114		70 - 130			03/20/23 11:02	04/11/23 18:21	1
13C2 PFHxA	104		70 - 130			03/20/23 11:02	04/11/23 18:21	1

Method: EPA 537 DW - Perfluorinated Alkyl Acids (LC/MS) - DL

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanoic acid	350		18	ng/L	03/20/23 11:02	04/17/23 12:29		10
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	100		70 - 130			03/20/23 11:02	04/17/23 12:29	10
13C2 PFDA	113		70 - 130			03/20/23 11:02	04/17/23 12:29	10
13C2 PFHxA	100		70 - 130			03/20/23 11:02	04/17/23 12:29	10

Client Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1

SDG: HOO

Client Sample ID: GAC Midfluent

Date Collected: 03/16/23 09:35
Date Received: 03/17/23 10:24

Lab Sample ID: 410-119243-2

Matrix: Water

Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L	04/07/23 09:19	04/14/23 18:48		1
8:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L	04/07/23 09:19	04/14/23 18:48		1
Perfluorobutanoic acid	4.6		1.7	ng/L	04/07/23 09:19	04/14/23 18:48		1
Perfluorodecanesulfonic acid	1.7	U	1.7	ng/L	04/07/23 09:19	04/14/23 18:48		1
Perfluoroheptanesulfonic acid	1.7	U	1.7	ng/L	04/07/23 09:19	04/14/23 18:48		1
Perfluoroctanesulfonamide	1.7	U	1.7	ng/L	04/07/23 09:19	04/14/23 18:48		1
Perfluoropentanoic acid	1.7	U	1.7	ng/L	04/07/23 09:19	04/14/23 18:48		1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
M2-6:2 FTS	121		17 - 200			04/07/23 09:19	04/14/23 18:48	1
M2-8:2 FTS	104		33 - 200			04/07/23 09:19	04/14/23 18:48	1
13C4 PFBA	110		42 - 165			04/07/23 09:19	04/14/23 18:48	1
13C5 PFPeA	106		38 - 187			04/07/23 09:19	04/14/23 18:48	1
13C8 PFOS	109		51 - 159			04/07/23 09:19	04/14/23 18:48	1
13C8 FOSA	91		10 - 168			04/07/23 09:19	04/14/23 18:48	1
13C3 PFHxA	111		28 - 188			04/07/23 09:19	04/14/23 18:48	1

Method: EPA 537 DW - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:31		1
Perfluoroheptanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:31		1
Perfluoroctanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:31		1
Perfluorononanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:31		1
Perfluorodecanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:31		1
Perfluorotridecanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:31		1
Perfluorotetradecanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:31		1
Perfluorobutanesulfonic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:31		1
Perfluorohexanesulfonic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:31		1
Perfluoroctanesulfonic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:31		1
NEtFOSAA	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:31		1
NMeFOSAA	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:31		1
Perfluoroundecanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:31		1
Perfluorododecanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:31		1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	92		70 - 130			03/20/23 11:02	04/11/23 18:31	1
13C2 PFDA	97		70 - 130			03/20/23 11:02	04/11/23 18:31	1
13C2 PFHxA	87		70 - 130			03/20/23 11:02	04/11/23 18:31	1

Client Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1

SDG: HOO

Client Sample ID: GAC Effluent

Date Collected: 03/16/23 09:40
Date Received: 03/17/23 10:24

Lab Sample ID: 410-119243-3

Matrix: Water

Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 22:57		1
8:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 22:57		1
Perfluorobutanoic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 22:57		1
Perfluorodecanesulfonic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 22:57		1
Perfluoroheptanesulfonic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 22:57		1
Perfluoroctanesulfonamide	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 22:57		1
Perfluoropentanoic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 22:57		1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
M2-6:2 FTS	124		17 - 200			03/28/23 08:13	04/07/23 22:57	1
M2-8:2 FTS	112		33 - 200			03/28/23 08:13	04/07/23 22:57	1
13C4 PFBA	96		42 - 165			03/28/23 08:13	04/07/23 22:57	1
13C5 PFPeA	97		38 - 187			03/28/23 08:13	04/07/23 22:57	1
13C8 PFOS	106		51 - 159			03/28/23 08:13	04/07/23 22:57	1
13C8 FOSA	84		10 - 168			03/28/23 08:13	04/07/23 22:57	1
13C3 PFHxA	116		28 - 188			03/28/23 08:13	04/07/23 22:57	1

Method: EPA 537 DW - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:42		1
Perfluoroheptanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:42		1
Perfluoroctanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:42		1
Perfluorononanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:42		1
Perfluorodecanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:42		1
Perfluorotridecanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:42		1
Perfluorotetradecanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:42		1
Perfluorobutanesulfonic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:42		1
Perfluorohexanesulfonic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:42		1
Perfluoroctanesulfonic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:42		1
NEtFOSAA	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:42		1
NMeFOSAA	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:42		1
Perfluoroundecanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:42		1
Perfluorododecanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:42		1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	94		70 - 130			03/20/23 11:02	04/11/23 18:42	1
13C2 PFDA	90		70 - 130			03/20/23 11:02	04/11/23 18:42	1
13C2 PFHxA	86		70 - 130			03/20/23 11:02	04/11/23 18:42	1

Client Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1

SDG: HOO

Client Sample ID: PV-1_25

Lab Sample ID: 410-119243-4

Matrix: Water

Date Collected: 03/16/23 09:43

Date Received: 03/17/23 10:24

Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:08		1
8:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:08		1
Perfluorobutanoic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:08		1
Perfluorodecanesulfonic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:08		1
Perfluoroheptanesulfonic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:08		1
Perfluoroctanesulfonamide	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:08		1
Perfluoropentanoic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:08		1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
M2-6:2 FTS	129		17 - 200			03/28/23 08:13	04/07/23 23:08	1
M2-8:2 FTS	116		33 - 200			03/28/23 08:13	04/07/23 23:08	1
13C4 PFBA	103		42 - 165			03/28/23 08:13	04/07/23 23:08	1
13C5 PFPeA	102		38 - 187			03/28/23 08:13	04/07/23 23:08	1
13C8 PFOS	104		51 - 159			03/28/23 08:13	04/07/23 23:08	1
13C8 FOSA	85		10 - 168			03/28/23 08:13	04/07/23 23:08	1
13C3 PFHxS	117		28 - 188			03/28/23 08:13	04/07/23 23:08	1

Method: EPA 537 DW - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:52		1
Perfluoroheptanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:52		1
Perfluoroctanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:52		1
Perfluorononanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:52		1
Perfluorodecanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:52		1
Perfluorotridecanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:52		1
Perfluorotetradecanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:52		1
Perfluorobutanesulfonic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:52		1
Perfluorohexanesulfonic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:52		1
Perfluoroctanesulfonic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:52		1
NEtFOSAA	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:52		1
NMeFOSAA	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:52		1
Perfluoroundecanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:52		1
Perfluorododecanoic acid	1.7	U	1.7	ng/L	03/20/23 11:02	04/11/23 18:52		1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	94		70 - 130			03/20/23 11:02	04/11/23 18:52	1
13C2 PFDA	94		70 - 130			03/20/23 11:02	04/11/23 18:52	1
13C2 PFHxA	90		70 - 130			03/20/23 11:02	04/11/23 18:52	1

Client Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1

SDG: HOO

Client Sample ID: PV-1_50
Date Collected: 03/16/23 09:45
Date Received: 03/17/23 10:24

Lab Sample ID: 410-119243-5
Matrix: Water

Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L	03/28/23 08:13	04/07/23 23:19		1
8:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L	03/28/23 08:13	04/07/23 23:19		1
Perfluorobutanoic acid	1.8	U	1.8	ng/L	03/28/23 08:13	04/07/23 23:19		1
Perfluorodecanesulfonic acid	1.8	U	1.8	ng/L	03/28/23 08:13	04/07/23 23:19		1
Perfluoroheptanesulfonic acid	1.8	U	1.8	ng/L	03/28/23 08:13	04/07/23 23:19		1
Perfluoroctanesulfonamide	1.8	U	1.8	ng/L	03/28/23 08:13	04/07/23 23:19		1
Perfluoropentanoic acid	1.8	U	1.8	ng/L	03/28/23 08:13	04/07/23 23:19		1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
M2-6:2 FTS	119		17 - 200			03/28/23 08:13	04/07/23 23:19	1
M2-8:2 FTS	102		33 - 200			03/28/23 08:13	04/07/23 23:19	1
13C4 PFBA	106		42 - 165			03/28/23 08:13	04/07/23 23:19	1
13C5 PFPeA	104		38 - 187			03/28/23 08:13	04/07/23 23:19	1
13C8 PFOS	109		51 - 159			03/28/23 08:13	04/07/23 23:19	1
13C8 FOSA	85		10 - 168			03/28/23 08:13	04/07/23 23:19	1
13C3 PFHxA	110		28 - 188			03/28/23 08:13	04/07/23 23:19	1

Method: EPA 537 DW - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:03		1
Perfluoroheptanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:03		1
Perfluoroctanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:03		1
Perfluorononanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:03		1
Perfluorodecanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:03		1
Perfluorotridecanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:03		1
Perfluorotetradecanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:03		1
Perfluorobutanesulfonic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:03		1
Perfluorohexanesulfonic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:03		1
Perfluoroctanesulfonic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:03		1
NEtFOSAA	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:03		1
NMeFOSAA	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:03		1
Perfluoroundecanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:03		1
Perfluorododecanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:03		1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	81		70 - 130			03/20/23 11:02	04/11/23 19:03	1
13C2 PFDA	94		70 - 130			03/20/23 11:02	04/11/23 19:03	1
13C2 PFHxA	88		70 - 130			03/20/23 11:02	04/11/23 19:03	1

Client Sample Results

Client: CT Male Associates DPC
 Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1

SDG: HOO

Client Sample ID: PV-1_75

Lab Sample ID: 410-119243-6

Matrix: Water

Date Collected: 03/16/23 09:47

Date Received: 03/17/23 10:24

Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:30		1
8:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:30		1
Perfluorobutanoic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:30		1
Perfluorodecanesulfonic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:30		1
Perfluoroheptanesulfonic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:30		1
Perfluoroctanesulfonamide	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:30		1
Perfluoropentanoic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:30		1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
M2-6:2 FTS	116		17 - 200			03/28/23 08:13	04/07/23 23:30	1
M2-8:2 FTS	103		33 - 200			03/28/23 08:13	04/07/23 23:30	1
13C4 PFBA	99		42 - 165			03/28/23 08:13	04/07/23 23:30	1
13C5 PFPeA	102		38 - 187			03/28/23 08:13	04/07/23 23:30	1
13C8 PFOS	106		51 - 159			03/28/23 08:13	04/07/23 23:30	1
13C8 FOSA	82		10 - 168			03/28/23 08:13	04/07/23 23:30	1
13C3 PFHxA	106		28 - 188			03/28/23 08:13	04/07/23 23:30	1

Method: EPA 537 DW - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:14		1
Perfluoroheptanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:14		1
Perfluoroctanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:14		1
Perfluorononanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:14		1
Perfluorodecanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:14		1
Perfluorotridecanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:14		1
Perfluorotetradecanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:14		1
Perfluorobutanesulfonic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:14		1
Perfluorohexanesulfonic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:14		1
Perfluoroctanesulfonic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:14		1
NEtFOSAA	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:14		1
NMeFOSAA	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:14		1
Perfluoroundecanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:14		1
Perfluorododecanoic acid	1.8	U	1.8	ng/L	03/20/23 11:02	04/11/23 19:14		1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	91		70 - 130			03/20/23 11:02	04/11/23 19:14	1
13C2 PFDA	97		70 - 130			03/20/23 11:02	04/11/23 19:14	1
13C2 PFHxA	85		70 - 130			03/20/23 11:02	04/11/23 19:14	1

Client Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1
SDG: HOO

Client Sample ID: FTB01_230316

Lab Sample ID: 410-119243-7

Matrix: Water

Date Collected: 03/16/23 09:55
Date Received: 03/17/23 10:24

Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:52		1
8:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:52		1
Perfluorobutanoic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:52		1
Perfluorodecanesulfonic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:52		1
Perfluoroheptanesulfonic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:52		1
Perfluoroctanesulfonamide	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:52		1
Perfluoropentanoic acid	1.7	U	1.7	ng/L	03/28/23 08:13	04/07/23 23:52		1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
M2-6:2 FTS	128		17 - 200			03/28/23 08:13	04/07/23 23:52	1
M2-8:2 FTS	112		33 - 200			03/28/23 08:13	04/07/23 23:52	1
13C4 PFBA	94		42 - 165			03/28/23 08:13	04/07/23 23:52	1
13C5 PFPeA	91		38 - 187			03/28/23 08:13	04/07/23 23:52	1
13C8 PFOS	105		51 - 159			03/28/23 08:13	04/07/23 23:52	1
13C8 FOSA	85		10 - 168			03/28/23 08:13	04/07/23 23:52	1
13C3 PFHxA	119		28 - 188			03/28/23 08:13	04/07/23 23:52	1

Method: EPA 537 DW - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	1.6	U cn	1.6	ng/L	03/20/23 11:02	04/11/23 19:24		1
Perfluoroheptanoic acid	1.6	U cn	1.6	ng/L	03/20/23 11:02	04/11/23 19:24		1
Perfluoroctanoic acid	1.6	U cn	1.6	ng/L	03/20/23 11:02	04/11/23 19:24		1
Perfluorononanoic acid	1.6	U cn	1.6	ng/L	03/20/23 11:02	04/11/23 19:24		1
Perfluorodecanoic acid	1.6	U cn	1.6	ng/L	03/20/23 11:02	04/11/23 19:24		1
Perfluorotridecanoic acid	1.6	U cn	1.6	ng/L	03/20/23 11:02	04/11/23 19:24		1
Perfluorotetradecanoic acid	1.6	U cn	1.6	ng/L	03/20/23 11:02	04/11/23 19:24		1
Perfluorobutanesulfonic acid	1.6	U cn	1.6	ng/L	03/20/23 11:02	04/11/23 19:24		1
Perfluorohexanesulfonic acid	1.6	U cn	1.6	ng/L	03/20/23 11:02	04/11/23 19:24		1
Perfluoroctanesulfonic acid	1.6	U cn	1.6	ng/L	03/20/23 11:02	04/11/23 19:24		1
NEtFOSAA	1.6	U cn	1.6	ng/L	03/20/23 11:02	04/11/23 19:24		1
NMeFOSAA	1.6	U cn	1.6	ng/L	03/20/23 11:02	04/11/23 19:24		1
Perfluoroundecanoic acid	1.6	U cn	1.6	ng/L	03/20/23 11:02	04/11/23 19:24		1
Perfluorododecanoic acid	1.6	U cn	1.6	ng/L	03/20/23 11:02	04/11/23 19:24		1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	48	S1- cn	70 - 130			03/20/23 11:02	04/11/23 19:24	1
13C2 PFDA	61	S1- cn	70 - 130			03/20/23 11:02	04/11/23 19:24	1
13C2 PFHxA	6	S1- cn	70 - 130			03/20/23 11:02	04/11/23 19:24	1

Client Sample Results

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1
SDG: HOO

Client Sample ID: LTB01_230316

Lab Sample ID: 410-119243-8

Date Collected: 03/16/23 00:00
Date Received: 03/17/23 10:24

Matrix: Water

Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L	03/28/23 08:13	04/08/23 00:03		1
8:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L	03/28/23 08:13	04/08/23 00:03		1
Perfluorobutanoic acid	1.8	U	1.8	ng/L	03/28/23 08:13	04/08/23 00:03		1
Perfluorodecanesulfonic acid	1.8	U	1.8	ng/L	03/28/23 08:13	04/08/23 00:03		1
Perfluoroheptanesulfonic acid	1.8	U	1.8	ng/L	03/28/23 08:13	04/08/23 00:03		1
Perfluoroctanesulfonamide	1.8	U	1.8	ng/L	03/28/23 08:13	04/08/23 00:03		1
Perfluoropentanoic acid	1.8	U	1.8	ng/L	03/28/23 08:13	04/08/23 00:03		1
Isotope Dilution	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
M2-6:2 FTS	119		17 - 200			03/28/23 08:13	04/08/23 00:03	1
M2-8:2 FTS	105		33 - 200			03/28/23 08:13	04/08/23 00:03	1
13C4 PFBA	105		42 - 165			03/28/23 08:13	04/08/23 00:03	1
13C5 PFPeA	102		38 - 187			03/28/23 08:13	04/08/23 00:03	1
13C8 PFOS	112		51 - 159			03/28/23 08:13	04/08/23 00:03	1
13C8 FOSA	92		10 - 168			03/28/23 08:13	04/08/23 00:03	1
13C3 PFHxA	110		28 - 188			03/28/23 08:13	04/08/23 00:03	1

Method: EPA 537 DW - Perfluorinated Alkyl Acids (LC/MS)

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorohexanoic acid	1.7	U cn	1.7	ng/L	03/20/23 11:02	04/11/23 19:35		1
Perfluoroheptanoic acid	1.7	U cn	1.7	ng/L	03/20/23 11:02	04/11/23 19:35		1
Perfluoroctanoic acid	1.7	U cn	1.7	ng/L	03/20/23 11:02	04/11/23 19:35		1
Perfluorononanoic acid	1.7	U cn	1.7	ng/L	03/20/23 11:02	04/11/23 19:35		1
Perfluorodecanoic acid	1.7	U cn	1.7	ng/L	03/20/23 11:02	04/11/23 19:35		1
Perfluorotridecanoic acid	1.7	U cn	1.7	ng/L	03/20/23 11:02	04/11/23 19:35		1
Perfluorotetradecanoic acid	1.7	U cn	1.7	ng/L	03/20/23 11:02	04/11/23 19:35		1
Perfluorobutanesulfonic acid	1.7	U cn	1.7	ng/L	03/20/23 11:02	04/11/23 19:35		1
Perfluorohexanesulfonic acid	1.7	U cn	1.7	ng/L	03/20/23 11:02	04/11/23 19:35		1
Perfluoroctanesulfonic acid	1.7	U cn	1.7	ng/L	03/20/23 11:02	04/11/23 19:35		1
NEtFOSAA	1.7	U cn	1.7	ng/L	03/20/23 11:02	04/11/23 19:35		1
NMeFOSAA	1.7	U cn	1.7	ng/L	03/20/23 11:02	04/11/23 19:35		1
Perfluoroundecanoic acid	1.7	U cn	1.7	ng/L	03/20/23 11:02	04/11/23 19:35		1
Perfluorododecanoic acid	1.7	U cn	1.7	ng/L	03/20/23 11:02	04/11/23 19:35		1
Surrogate	%Recovery	Qualifier	Limits			Prepared	Analyzed	Dil Fac
d5-NEtFOSAA	46	S1- cn	70 - 130			03/20/23 11:02	04/11/23 19:35	1
13C2 PFDA	75	cn	70 - 130			03/20/23 11:02	04/11/23 19:35	1
13C2 PFHxA	16	S1- cn	70 - 130			03/20/23 11:02	04/11/23 19:35	1

Surrogate Summary

Client: CT Male Associates DPC
 Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1
 SDG: HOO

Method: 537 DW - Perfluorinated Alkyl Acids (LC/MS)

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		d5NEFOS (70-130)	PFDA (70-130)	PFHxA (70-130)
410-119243-1	GAC Influent	91	114	104
410-119243-1 - DL	GAC Influent	100	113	100
410-119243-2	GAC Midfluent	92	97	87
410-119243-3	GAC Effluent	94	90	86
410-119243-4	PV-1_25	94	94	90
410-119243-5	PV-1_50	81	94	88
410-119243-6	PV-1_75	91	97	85
410-119243-7	FTB01_230316	48 S1- cn	61 S1- cn	6 S1- cn
410-119243-8	LTB01_230316	46 S1- cn	75 cn	16 S1- cn
LCS 410-355279/14-A	Lab Control Sample	97	94	87
LCSD 410-355279/15-A	Lab Control Sample Dup	90	89	83
MB 410-355279/13-A	Method Blank	92	87	79

Surrogate Legend

d5NEFOS = d5-NEtFOSAA

PFDA = 13C2 PFDA

PFHxA = 13C2 PFHxA

Isotope Dilution Summary

Client: CT Male Associates DPC
 Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1
 SDG: HOO

Method: 537 (Mod) - EPA 537 Version 1.1 modified

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Isotope Dilution Recovery (Acceptance Limits)						
		M262FTS (17-200)	M282FTS (33-200)	PFBA (42-165)	PFPeA (38-187)	C8PFOS (51-159)	PFOSA (10-168)	C3PFHS (28-188)
410-119243-1	GAC Influent	106	134	118	123	115	92	135
410-119243-2	GAC Midfluent	121	104	110	106	109	91	111
410-119243-3	GAC Effluent	124	112	96	97	106	84	116
410-119243-4	PV-1_25	129	116	103	102	104	85	117
410-119243-5	PV-1_50	119	102	106	104	109	85	110
410-119243-6	PV-1_75	116	103	99	102	106	82	106
410-119243-7	FTB01_230316	128	112	94	91	105	85	119
410-119243-8	LTB01_230316	119	105	105	102	112	92	110
LCS 410-358051/3-A	Lab Control Sample	119	110	109	110	108	95	109
LCS 410-361976/2-A	Lab Control Sample	116	117	111	110	114	100	107
MB 410-358051/1-A	Method Blank	120	107	105	109	106	88	112
MB 410-361976/1-A	Method Blank	113	104	104	102	105	94	118
Surrogate Legend								
M262FTS = M2-6:2 FTS								
M282FTS = M2-8:2 FTS								
PFBA = 13C4 PFBA								
PFPeA = 13C5 PFPeA								
C8PFOS = 13C8 PFOS								
PFOSA = 13C8 FOSA								
C3PFHS = 13C3 PFHxS								

QC Sample Results

Client: CT Male Associates DPC
 Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1
 SDG: HOO

Method: 537 (Mod) - EPA 537 Version 1.1 modified

Lab Sample ID: MB 410-358051/1-A

Matrix: Water

Analysis Batch: 362222

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 358051

Analyte	MB	MB	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier								
6:2 Fluorotelomer sulfonic acid	2.0	U	2.0		2.0	ng/L	03/28/23 08:13	04/07/23 19:14		1
8:2 Fluorotelomer sulfonic acid	2.0	U	2.0		2.0	ng/L	03/28/23 08:13	04/07/23 19:14		1
Perfluorobutanoic acid	2.0	U	2.0		2.0	ng/L	03/28/23 08:13	04/07/23 19:14		1
Perfluorodecanesulfonic acid	2.0	U	2.0		2.0	ng/L	03/28/23 08:13	04/07/23 19:14		1
Perfluoroheptanesulfonic acid	2.0	U	2.0		2.0	ng/L	03/28/23 08:13	04/07/23 19:14		1
Perfluorooctanesulfonamide	2.0	U	2.0		2.0	ng/L	03/28/23 08:13	04/07/23 19:14		1
Perfluoropentanoic acid	2.0	U	2.0		2.0	ng/L	03/28/23 08:13	04/07/23 19:14		1
<i>Isotope Dilution</i>	MB	MB	%Recovery	Qualifier	Limits		D	Prepared	Analyzed	Dil Fac
	Result	Qualifier								
M2-6:2 FTS	120		17 - 200				03/28/23 08:13	04/07/23 19:14		1
M2-8:2 FTS	107		33 - 200				03/28/23 08:13	04/07/23 19:14		1
13C4 PFBA	105		42 - 165				03/28/23 08:13	04/07/23 19:14		1
13C5 PFPeA	109		38 - 187				03/28/23 08:13	04/07/23 19:14		1
13C8 PFOS	106		51 - 159				03/28/23 08:13	04/07/23 19:14		1
13C8 FOSA	88		10 - 168				03/28/23 08:13	04/07/23 19:14		1
13C3 PFHxS	112		28 - 188				03/28/23 08:13	04/07/23 19:14		1

Lab Sample ID: LCS 410-358051/3-A

Matrix: Water

Analysis Batch: 362222

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 358051

Analyte	Spike	LCS	LCS	Result	Qualifier	Unit	D	%Rec	Limits	
	Added	Result	Qualifier							
6:2 Fluorotelomer sulfonic acid	24.3	20.1		ng/L	83	28 - 173				
8:2 Fluorotelomer sulfonic acid	24.5	22.2		ng/L	91	55 - 138				
Perfluorobutanoic acid	25.6	20.2		ng/L	79	59 - 136				
Perfluorodecanesulfonic acid	24.7	20.1		ng/L	81	55 - 137				
Perfluoroheptanesulfonic acid	24.4	18.7		ng/L	77	56 - 140				
Perfluorooctanesulfonamide	25.6	22.8		ng/L	89	43 - 167				
Perfluoropentanoic acid	25.6	20.5		ng/L	80	57 - 141				
<i>Isotope Dilution</i>	LCS	LCS	%Recovery	Qualifier	Limits		D	%Rec	Limits	
	Result	Qualifier								
M2-6:2 FTS	119	17 - 200								
M2-8:2 FTS	110	33 - 200								
13C4 PFBA	109	42 - 165								
13C5 PFPeA	110	38 - 187								
13C8 PFOS	108	51 - 159								
13C8 FOSA	95	10 - 168								
13C3 PFHxS	109	28 - 188								

Lab Sample ID: MB 410-361976/1-A

Matrix: Water

Analysis Batch: 364632

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 361976

Analyte	MB	MB	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier								
6:2 Fluorotelomer sulfonic acid	2.0	U	2.0		2.0	ng/L	04/07/23 09:19	04/14/23 17:53		1
8:2 Fluorotelomer sulfonic acid	2.0	U	2.0		2.0	ng/L	04/07/23 09:19	04/14/23 17:53		1
Perfluorobutanoic acid	2.0	U	2.0		2.0	ng/L	04/07/23 09:19	04/14/23 17:53		1
Perfluorodecanesulfonic acid	2.0	U	2.0		2.0	ng/L	04/07/23 09:19	04/14/23 17:53		1
Perfluoroheptanesulfonic acid	2.0	U	2.0		2.0	ng/L	04/07/23 09:19	04/14/23 17:53		1

Eurofins Lancaster Laboratories Environment Testing, LLC

QC Sample Results

Client: CT Male Associates DPC
 Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1
 SDG: HOO

Method: 537 (Mod) - EPA 537 Version 1.1 modified (Continued)

Lab Sample ID: MB 410-361976/1-A

Client Sample ID: Method Blank

Matrix: Water

Prep Type: Total/NA

Analysis Batch: 364632

Prep Batch: 361976

Analyte	MB		RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
Perfluoroctanesulfonamide	2.0	U	2.0	ng/L	04/07/23 09:19	04/14/23 17:53		1
Perfluoropentanoic acid	2.0	U	2.0	ng/L	04/07/23 09:19	04/14/23 17:53		1
Isotope Dilution								
M2-6:2 FTS	113		17 - 200		04/07/23 09:19	04/14/23 17:53		1
M2-8:2 FTS	104		33 - 200		04/07/23 09:19	04/14/23 17:53		1
13C4 PFBA	104		42 - 165		04/07/23 09:19	04/14/23 17:53		1
13C5 PPPeA	102		38 - 187		04/07/23 09:19	04/14/23 17:53		1
13C8 PFOS	105		51 - 159		04/07/23 09:19	04/14/23 17:53		1
13C8 FOSA	94		10 - 168		04/07/23 09:19	04/14/23 17:53		1
13C3 PFHxS	118		28 - 188		04/07/23 09:19	04/14/23 17:53		1

Lab Sample ID: LCS 410-361976/2-A

Client Sample ID: Lab Control Sample

Matrix: Water

Prep Type: Total/NA

Analysis Batch: 364632

Prep Batch: 361976

Analyte	Spike		Result	Qualifier	Unit	D	%Rec	Limits
	Added							
6:2 Fluorotelomer sulfonic acid	24.3		26.4	ng/L	109	28 - 173		
8:2 Fluorotelomer sulfonic acid	24.5		25.6	ng/L	104	55 - 138		
Perfluorobutanoic acid	25.6		26.6	ng/L	104	59 - 136		
Perfluorodecanesulfonic acid	24.7		25.9	ng/L	105	55 - 137		
Perfluoroheptanesulfonic acid	24.4		26.6	ng/L	109	56 - 140		
Perfluoroctanesulfonamide	25.6		30.1	ng/L	118	43 - 167		
Perfluoropentanoic acid	25.6		28.1	ng/L	110	57 - 141		
Isotope Dilution								
M2-6:2 FTS	116		17 - 200					
M2-8:2 FTS	117		33 - 200					
13C4 PFBA	111		42 - 165					
13C5 PPPeA	110		38 - 187					
13C8 PFOS	114		51 - 159					
13C8 FOSA	100		10 - 168					
13C3 PFHxS	107		28 - 188					

Method: 537 DW - Perfluorinated Alkyl Acids (LC/MS)

Lab Sample ID: MB 410-355279/13-A

Client Sample ID: Method Blank

Matrix: Water

Prep Type: Total/NA

Analysis Batch: 363162

Prep Batch: 355279

Analyte	MB		RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
Perfluorohexanoic acid	2.0	U	2.0	ng/L	03/20/23 11:02	04/11/23 17:38		1
Perfluoroheptanoic acid	2.0	U	2.0	ng/L	03/20/23 11:02	04/11/23 17:38		1
Perfluoroctanoic acid	2.0	U	2.0	ng/L	03/20/23 11:02	04/11/23 17:38		1
Perfluorononanoic acid	2.0	U	2.0	ng/L	03/20/23 11:02	04/11/23 17:38		1
Perfluorodecanoic acid	2.0	U	2.0	ng/L	03/20/23 11:02	04/11/23 17:38		1
Perfluorotridecanoic acid	2.0	U	2.0	ng/L	03/20/23 11:02	04/11/23 17:38		1
Perfluorotetradecanoic acid	2.0	U	2.0	ng/L	03/20/23 11:02	04/11/23 17:38		1
Perfluorobutanesulfonic acid	2.0	U	2.0	ng/L	03/20/23 11:02	04/11/23 17:38		1

QC Sample Results

Client: CT Male Associates DPC
 Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1
 SDG: HOO

Method: 537 DW - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: MB 410-355279/13-A

Matrix: Water

Analysis Batch: 363162

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 355279

Analyte	MB	MB	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
	Result	Qualifier								
Perfluorohexanesulfonic acid	2.0	U	2.0		2.0	ng/L	03/20/23 11:02	04/11/23 17:38		1
Perfluoroctanesulfonic acid	2.0	U	2.0		2.0	ng/L	03/20/23 11:02	04/11/23 17:38		1
NEtFOSAA	2.0	U	2.0		2.0	ng/L	03/20/23 11:02	04/11/23 17:38		1
NMeFOSAA	2.0	U	2.0		2.0	ng/L	03/20/23 11:02	04/11/23 17:38		1
Perfluoroundecanoic acid	2.0	U	2.0		2.0	ng/L	03/20/23 11:02	04/11/23 17:38		1
Perfluorododecanoic acid	2.0	U	2.0		2.0	ng/L	03/20/23 11:02	04/11/23 17:38		1

Surrogate	MB	MB	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
	Result	Qualifier						
d5-NEtFOSAA	92		70 - 130			03/20/23 11:02	04/11/23 17:38	1
13C2 PFDA	87		70 - 130			03/20/23 11:02	04/11/23 17:38	1
13C2 PFHxA	79		70 - 130			03/20/23 11:02	04/11/23 17:38	1

Lab Sample ID: LCS 410-355279/14-A

Matrix: Water

Analysis Batch: 363162

Client Sample ID: Lab Control Sample
 Prep Type: Total/NA
 Prep Batch: 355279

Analyte	Spike	LCS	LCS	Result	Qualifier	Unit	D	%Rec	Limits	%Rec
	Added	Result	Qualifier							
Perfluorohexanoic acid	20.5	17.5		17.5		ng/L		85	70 - 130	
Perfluoroheptanoic acid	20.5	17.5		17.5		ng/L		85	70 - 130	
Perfluoroctanoic acid	20.5	17.5		17.5		ng/L		85	70 - 130	
Perfluorononanoic acid	20.5	17.8		17.8		ng/L		87	70 - 130	
Perfluorodecanoic acid	20.5	19.0		19.0		ng/L		93	70 - 130	
Perfluorotridecanoic acid	20.5	16.1		16.1		ng/L		78	70 - 130	
Perfluorotetradecanoic acid	20.5	18.7		18.7		ng/L		92	70 - 130	
Perfluorobutanesulfonic acid	18.1	14.6		14.6		ng/L		81	70 - 130	
Perfluorohexanesulfonic acid	18.7	15.0		15.0		ng/L		80	70 - 130	
Perfluoroctanesulfonic acid	19.0	15.8		15.8		ng/L		83	70 - 130	
NEtFOSAA	20.5	17.7		17.7		ng/L		86	70 - 130	
NMeFOSAA	20.5	17.3		17.3		ng/L		84	70 - 130	
Perfluoroundecanoic acid	20.5	17.2		17.2		ng/L		84	70 - 130	
Perfluorododecanoic acid	20.5	16.0		16.0		ng/L		78	70 - 130	

Surrogate	LCS	LCS	%Recovery	Qualifier	Limits
	Added	Result			
d5-NEtFOSAA	97	70 - 130			
13C2 PFDA	94	70 - 130			
13C2 PFHxA	87	70 - 130			

Lab Sample ID: LCSD 410-355279/15-A

Matrix: Water

Analysis Batch: 363162

Client Sample ID: Lab Control Sample Dup
 Prep Type: Total/NA
 Prep Batch: 355279

Analyte	Spike	LCSD	LCSD	Result	Qualifier	Unit	D	%Rec	Limits	RPD	Limit
	Added	Result	Qualifier								
Perfluorohexanoic acid	20.5	17.0		17.0		ng/L		83	70 - 130	3	30
Perfluoroheptanoic acid	20.5	17.7		17.7		ng/L		87	70 - 130	1	30
Perfluoroctanoic acid	20.5	18.2		18.2		ng/L		89	70 - 130	4	30
Perfluorononanoic acid	20.5	17.6		17.6		ng/L		86	70 - 130	1	30
Perfluorodecanoic acid	20.5	18.5		18.5		ng/L		90	70 - 130	3	30
Perfluorotridecanoic acid	20.5	16.7		16.7		ng/L		81	70 - 130	4	30

Eurofins Lancaster Laboratories Environment Testing, LLC

QC Sample Results

Client: CT Male Associates DPC
 Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1
 SDG: HOO

Method: 537 DW - Perfluorinated Alkyl Acids (LC/MS) (Continued)

Lab Sample ID: LCSD 410-355279/15-A

Matrix: Water

Analysis Batch: 363162

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 355279

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	Limits	RPD	RPD Limit
Perfluorotetradecanoic acid	20.5	20.1		ng/L	98	70 - 130	7	30	
Perfluorobutanesulfonic acid	18.1	13.9		ng/L	77	70 - 130	5	30	
Perfluorohexamersulfonic acid	18.7	15.0		ng/L	80	70 - 130	0	30	
Perfluoroctanesulfonic acid	19.0	15.9		ng/L	84	70 - 130	1	30	
NEtFOSAA	20.5	16.8		ng/L	82	70 - 130	5	30	
NMeFOSAA	20.5	15.8		ng/L	77	70 - 130	9	30	
Perfluoroundecanoic acid	20.5	17.7		ng/L	86	70 - 130	3	30	
Perfluorododecanoic acid	20.5	16.5		ng/L	80	70 - 130	3	30	

Surrogate	LCSD	LCSD	Limits
	%Recovery	Qualifier	
d5-NEtFOSAA	90		70 - 130
13C2 PFDA	89		70 - 130
13C2 PFHxA	83		70 - 130

QC Association Summary

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1
SDG: HOO

LCMS

Prep Batch: 355279

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-119243-1	GAC Influent	Total/NA	Water	537 DW	1
410-119243-1 - DL	GAC Influent	Total/NA	Water	537 DW	2
410-119243-2	GAC Midfluent	Total/NA	Water	537 DW	3
410-119243-3	GAC Effluent	Total/NA	Water	537 DW	4
410-119243-4	PV-1_25	Total/NA	Water	537 DW	5
410-119243-5	PV-1_50	Total/NA	Water	537 DW	6
410-119243-6	PV-1_75	Total/NA	Water	537 DW	7
410-119243-7	FTB01_230316	Total/NA	Water	537 DW	8
410-119243-8	LTB01_230316	Total/NA	Water	537 DW	9
MB 410-355279/13-A	Method Blank	Total/NA	Water	537 DW	10
LCS 410-355279/14-A	Lab Control Sample	Total/NA	Water	537 DW	
LCSD 410-355279/15-A	Lab Control Sample Dup	Total/NA	Water	537 DW	

Prep Batch: 358051

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-119243-3	GAC Effluent	Total/NA	Water	SPE	11
410-119243-4	PV-1_25	Total/NA	Water	SPE	12
410-119243-5	PV-1_50	Total/NA	Water	SPE	13
410-119243-6	PV-1_75	Total/NA	Water	SPE	14
410-119243-7	FTB01_230316	Total/NA	Water	SPE	15
410-119243-8	LTB01_230316	Total/NA	Water	SPE	
MB 410-358051/1-A	Method Blank	Total/NA	Water	SPE	
LCS 410-358051/3-A	Lab Control Sample	Total/NA	Water	SPE	

Prep Batch: 361976

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-119243-1	GAC Influent	Total/NA	Water	SPE	
410-119243-2	GAC Midfluent	Total/NA	Water	SPE	
MB 410-361976/1-A	Method Blank	Total/NA	Water	SPE	
LCS 410-361976/2-A	Lab Control Sample	Total/NA	Water	SPE	

Analysis Batch: 362222

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-119243-3	GAC Effluent	Total/NA	Water	537 (Mod)	358051
410-119243-4	PV-1_25	Total/NA	Water	537 (Mod)	358051
410-119243-5	PV-1_50	Total/NA	Water	537 (Mod)	358051
410-119243-6	PV-1_75	Total/NA	Water	537 (Mod)	358051
410-119243-7	FTB01_230316	Total/NA	Water	537 (Mod)	358051
410-119243-8	LTB01_230316	Total/NA	Water	537 (Mod)	358051
MB 410-358051/1-A	Method Blank	Total/NA	Water	537 (Mod)	358051
LCS 410-358051/3-A	Lab Control Sample	Total/NA	Water	537 (Mod)	358051

Analysis Batch: 363162

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-119243-1	GAC Influent	Total/NA	Water	537 DW	355279
410-119243-2	GAC Midfluent	Total/NA	Water	537 DW	355279
410-119243-3	GAC Effluent	Total/NA	Water	537 DW	355279
410-119243-4	PV-1_25	Total/NA	Water	537 DW	355279
410-119243-5	PV-1_50	Total/NA	Water	537 DW	355279
410-119243-6	PV-1_75	Total/NA	Water	537 DW	355279
410-119243-7	FTB01_230316	Total/NA	Water	537 DW	355279

QC Association Summary

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1
SDG: HOO

LCMS (Continued)

Analysis Batch: 363162 (Continued)

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-119243-8	LTB01_230316	Total/NA	Water	537 DW	355279
MB 410-355279/13-A	Method Blank	Total/NA	Water	537 DW	355279
LCS 410-355279/14-A	Lab Control Sample	Total/NA	Water	537 DW	355279
LCSD 410-355279/15-A	Lab Control Sample Dup	Total/NA	Water	537 DW	355279

Analysis Batch: 364632

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-119243-1	GAC Influent	Total/NA	Water	537 (Mod)	361976
410-119243-2	GAC Midfluent	Total/NA	Water	537 (Mod)	361976
MB 410-361976/1-A	Method Blank	Total/NA	Water	537 (Mod)	361976
LCS 410-361976/2-A	Lab Control Sample	Total/NA	Water	537 (Mod)	361976

Analysis Batch: 365037

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-119243-1 - DL	GAC Influent	Total/NA	Water	537 DW	355279

Prep Batch: 365743

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-119243-7 - RE	FTB01_230316	Total/NA	Water	537 DW	13
410-119243-8 - RE	LTB01_230316	Total/NA	Water	537 DW	14
MB 410-365743/1-A	Method Blank	Total/NA	Water	537 DW	15
LCS 410-365743/2-A	Lab Control Sample	Total/NA	Water	537 DW	16
LCSD 410-365743/3-A	Lab Control Sample Dup	Total/NA	Water	537 DW	
LLCS 410-365743/4-A	Lab Control Sample	Total/NA	Water	537 DW	

Analysis Batch: 367424

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-119243-7 - RE	FTB01_230316	Total/NA	Water	537 DW	365743
410-119243-8 - RE	LTB01_230316	Total/NA	Water	537 DW	365743
MB 410-365743/1-A	Method Blank	Total/NA	Water	537 DW	365743
LCS 410-365743/2-A	Lab Control Sample	Total/NA	Water	537 DW	365743
LCSD 410-365743/3-A	Lab Control Sample Dup	Total/NA	Water	537 DW	365743
LLCS 410-365743/4-A	Lab Control Sample	Total/NA	Water	537 DW	365743

Lab Chronicle

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1

SDG: HOO

Client Sample ID: GAC Influent

Date Collected: 03/16/23 09:30

Date Received: 03/17/23 10:24

Lab Sample ID: 410-119243-1

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			361976	S7AC	ELLE	04/07/23 09:19
Total/NA	Analysis	537 (Mod)		1	364632	DTA4	ELLE	04/14/23 18:37
Total/NA	Prep	537 DW			355279	HQ8B	ELLE	03/20/23 11:02
Total/NA	Analysis	537 DW		1	363162	TAS6	ELLE	04/11/23 18:21
Total/NA	Prep	537 DW	DL		355279	HQ8B	ELLE	03/20/23 11:02
Total/NA	Analysis	537 DW	DL	10	365037	TAS6	ELLE	04/17/23 12:29

Client Sample ID: GAC Midfluent

Date Collected: 03/16/23 09:35

Date Received: 03/17/23 10:24

Lab Sample ID: 410-119243-2

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			361976	S7AC	ELLE	04/07/23 09:19
Total/NA	Analysis	537 (Mod)		1	364632	DTA4	ELLE	04/14/23 18:48
Total/NA	Prep	537 DW			355279	HQ8B	ELLE	03/20/23 11:02
Total/NA	Analysis	537 DW		1	363162	TAS6	ELLE	04/11/23 18:31

Client Sample ID: GAC Effluent

Date Collected: 03/16/23 09:40

Date Received: 03/17/23 10:24

Lab Sample ID: 410-119243-3

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			358051	D5VP	ELLE	03/28/23 08:13
Total/NA	Analysis	537 (Mod)		1	362222	VK3G	ELLE	04/07/23 22:57
Total/NA	Prep	537 DW			355279	HQ8B	ELLE	03/20/23 11:02
Total/NA	Analysis	537 DW		1	363162	TAS6	ELLE	04/11/23 18:42

Client Sample ID: PV-1_25

Date Collected: 03/16/23 09:43

Date Received: 03/17/23 10:24

Lab Sample ID: 410-119243-4

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			358051	D5VP	ELLE	03/28/23 08:13
Total/NA	Analysis	537 (Mod)		1	362222	VK3G	ELLE	04/07/23 23:08
Total/NA	Prep	537 DW			355279	HQ8B	ELLE	03/20/23 11:02
Total/NA	Analysis	537 DW		1	363162	TAS6	ELLE	04/11/23 18:52

Client Sample ID: PV-1_50

Date Collected: 03/16/23 09:45

Date Received: 03/17/23 10:24

Lab Sample ID: 410-119243-5

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			358051	D5VP	ELLE	03/28/23 08:13
Total/NA	Analysis	537 (Mod)		1	362222	VK3G	ELLE	04/07/23 23:19
Total/NA	Prep	537 DW			355279	HQ8B	ELLE	03/20/23 11:02
Total/NA	Analysis	537 DW		1	363162	TAS6	ELLE	04/11/23 19:03

Lab Chronicle

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1

SDG: HOO

Client Sample ID: PV-1_75

Date Collected: 03/16/23 09:47

Date Received: 03/17/23 10:24

Lab Sample ID: 410-119243-6

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			358051	D5VP	ELLE	03/28/23 08:13
Total/NA	Analysis	537 (Mod)		1	362222	VK3G	ELLE	04/07/23 23:30
Total/NA	Prep	537 DW			355279	HQ8B	ELLE	03/20/23 11:02
Total/NA	Analysis	537 DW		1	363162	TAS6	ELLE	04/11/23 19:14

Client Sample ID: FTB01_230316

Date Collected: 03/16/23 09:55

Date Received: 03/17/23 10:24

Lab Sample ID: 410-119243-7

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			358051	D5VP	ELLE	03/28/23 08:13
Total/NA	Analysis	537 (Mod)		1	362222	VK3G	ELLE	04/07/23 23:52
Total/NA	Prep	537 DW			355279	HQ8B	ELLE	03/20/23 11:02
Total/NA	Analysis	537 DW		1	363162	TAS6	ELLE	04/11/23 19:24
Total/NA	Prep	537 DW	RE		365743	WW2J	ELLE	04/18/23 17:44
Total/NA	Analysis	537 DW	RE	1	367424	VK3G	ELLE	04/23/23 00:56

Client Sample ID: LTB01_230316

Date Collected: 03/16/23 00:00

Date Received: 03/17/23 10:24

Lab Sample ID: 410-119243-8

Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			358051	D5VP	ELLE	03/28/23 08:13
Total/NA	Analysis	537 (Mod)		1	362222	VK3G	ELLE	04/08/23 00:03
Total/NA	Prep	537 DW			355279	HQ8B	ELLE	03/20/23 11:02
Total/NA	Analysis	537 DW		1	363162	TAS6	ELLE	04/11/23 19:35
Total/NA	Prep	537 DW	RE		365743	WW2J	ELLE	04/18/23 17:44
Total/NA	Analysis	537 DW	RE	1	367424	VK3G	ELLE	04/23/23 01:06

Laboratory References:

ELLE = Eurofins Lancaster Laboratories Environment Testing, LLC, 2425 New Holland Pike, Lancaster, PA 17601, TEL (717)656-2300

Accreditation/Certification Summary

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1
SDG: HOO

Laboratory: Eurofins Lancaster Laboratories Environment Testing, LLC

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
New York	NELAP	10670	04-01-24

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
537 (Mod)	SPE	Water	6:2 Fluorotelomer sulfonic acid
537 (Mod)	SPE	Water	8:2 Fluorotelomer sulfonic acid
537 (Mod)	SPE	Water	Perfluorobutanoic acid
537 (Mod)	SPE	Water	Perfluorodecanesulfonic acid
537 (Mod)	SPE	Water	Perfluoroheptanesulfonic acid
537 (Mod)	SPE	Water	Perfluoroctanesulfonamide
537 (Mod)	SPE	Water	Perfluoropentanoic acid
537 DW	537 DW	Water	NEtFOSAA
537 DW	537 DW	Water	NMeFOSAA
537 DW	537 DW	Water	Perfluorobutanesulfonic acid
537 DW	537 DW	Water	Perfluorodecanoic acid
537 DW	537 DW	Water	Perfluorododecanoic acid
537 DW	537 DW	Water	Perfluoroheptanoic acid
537 DW	537 DW	Water	Perfluorohexanesulfonic acid
537 DW	537 DW	Water	Perfluorohexanoic acid
537 DW	537 DW	Water	Perfluorononanoic acid
537 DW	537 DW	Water	Perfluoroctanesulfonic acid
537 DW	537 DW	Water	Perfluorooctanoic acid
537 DW	537 DW	Water	Perfluorotetradecanoic acid
537 DW	537 DW	Water	Perfluorotridecanoic acid
537 DW	537 DW	Water	Perfluoroundecanoic acid

Method Summary

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1

SDG: HOO

Method	Method Description	Protocol	Laboratory
537 (Mod)	EPA 537 Version 1.1 modified	EPA	ELLE
537 DW	Perfluorinated Alkyl Acids (LC/MS)	EPA	ELLE
537 DW	Extraction of Perfluorinated Alkyl Acids	EPA	ELLE

Protocol References:

EPA = US Environmental Protection Agency

Laboratory References:

ELLE = Eurofins Lancaster Laboratories Environment Testing, LLC, 2425 New Holland Pike, Lancaster, PA 17601, TEL (717)656-2300

1

2

3

4

5

6

7

8

9

10

11

12

13

14

15

16

Sample Summary

Client: CT Male Associates DPC
Project/Site: Hoosick Falls WTP

Job ID: 410-119243-1
SDG: HOO

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
410-119243-1	GAC Influent	Water	03/16/23 09:30	03/17/23 10:24
410-119243-2	GAC Midfluent	Water	03/16/23 09:35	03/17/23 10:24
410-119243-3	GAC Effluent	Water	03/16/23 09:40	03/17/23 10:24
410-119243-4	PV-1_25	Water	03/16/23 09:43	03/17/23 10:24
410-119243-5	PV-1_50	Water	03/16/23 09:45	03/17/23 10:24
410-119243-6	PV-1_75	Water	03/16/23 09:47	03/17/23 10:24
410-119243-7	FTB01_230316	Water	03/16/23 09:55	03/17/23 10:24
410-119243-8	LTB01_230316	Water	03/16/23 00:00	03/17/23 10:24



1vironme

Chain of Custody Record

 eurofins
Environment Testing

410-119243 Chain of Custody

Client Contact: Jonathan Dippert, Kirk Malone		Sampler: C. Ormsby	Lab PM: Hobart, Paul	Carmer Tracking No(s):	COC No: 410-42503-12960.2		
Company: CT Male Associates DPC		Phone:	E-Mail: Paul.Hobart@et.eurofinsus.com	State of Origin: NV	Page: 2022 1 of 1		
Address: 50 Century Hill Dr		PWSID:	Analysis Requested				
City: Latham		Due Date Requested:					
State, Zip: NY, 12110		TAT Requested (days): <i>Standars</i>					
Phone:		Compliance Project: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No					
Email: j.dippert@ctmale.com, K. Malone@ctmale.com		PO #:					
Project Name: Hoosick Falls WTP		Purchase Order not required					
Site:		WO #:					
SSOW#:							
Sample Identification		Sample Date	Sample Time	Sample Type (C=comp, G=grab)	Matrix (W=water, S=solid, O=waste/oil, BT=tissue, A=air)		
					Field Filtered Sample (Yes or No)		
					Perform MS/MSD (Yes or No)		
					PFC_IDA - (MOD) 7 PFAS Compounds		
					537_DW - 14 PFAS Drinking Water List		
					537_DW - 14 PFAS Drinking Water List		
					Total Number of containers		
					Special Instructions/Note: <i>PFAS QC Batch collection here</i>		
<i>GAC INFLUENT</i>		<i>3/16/23</i>	<i>0930</i>	<i>G</i>	Water	<i>N N X Y</i>	<i>8</i>
<i>GAC MIDFLUENT</i>			<i>0935</i>		Water	<i>X Y</i>	<i>4</i>
<i>GAC EFFLUENT</i>			<i>0940</i>		Water	<i>X Y</i>	<i>4</i>
<i>PV-L_25</i>			<i>0941</i>			<i>X Y</i>	<i>4</i>
<i>PV-L_50</i>			<i>0945</i>			<i>X Y</i>	<i>4</i>
<i>PV-L_75</i>			<i>0947</i>			<i>XX</i>	<i>4</i>
<i>FTB01-230316</i>			<i>0955</i>			<i>XX</i>	<i>4</i>
<i>LTB01-230316</i>			<i>-</i>			<i>V X X</i>	<i>4</i>
Possible Hazard Identification		Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)					
<input type="checkbox"/> Non-Hazard <input type="checkbox"/> Flammable <input type="checkbox"/> Skin Irritant <input type="checkbox"/> Poison B <input type="checkbox"/> Unknown <input type="checkbox"/> Radiological		<input type="checkbox"/> Return To Client <input type="checkbox"/> Disposal By Lab <input type="checkbox"/> Archive For _____ Months					
Deliverable Requested: I, II, III, IV, Other (specify): <i>ASP-B, EQUIS 1 FILE</i>		Special Instructions/QC Requirements:					
Empty Kit Relinquished by: <i>Christina Ormsby</i>		Date: <i>3/16/23</i>	Time: <i>110</i>	Method of Shipment:			
Relinquished by: <i>Christina Ormsby</i>	Date/Time: <i>3/16/23 110</i>	Company: <i>CM</i>	Received by: <i>MMF</i>	Date/Time: <i>3/17/23 1024</i>	Company: <i>MNF</i>		
Relinquished by: <i>Christina Ormsby</i>	Date/Time: <i>3/16/23 110</i>	Company: <i>CM</i>	Received by: <i>MMF</i>	Date/Time: <i>3/17/23 1024</i>	Company: <i>MNF</i>		
Relinquished by: <i>Christina Ormsby</i>	Date/Time: <i>3/16/23 110</i>	Company: <i>CM</i>	Received by: <i>MMF</i>	Date/Time: <i>3/17/23 1024</i>	Company: <i>MNF</i>		
Custody Seals Intact: <input checked="" type="checkbox"/> Yes <input type="checkbox"/> No		Custody Seal No.: <i>3.9</i>		Cooler Temperature(s) °C and Other Remarks:			

Login Sample Receipt Checklist

Client: CT Male Associates DPC

Job Number: 410-119243-1

SDG Number: HOO

Login Number: 119243

List Source: Eurofins Lancaster Laboratories Environment Testing, LLC

List Number: 1

Creator: Reiff, Nicole L

Question	Answer	Comment	
The cooler's custody seal is intact.	True		1
The cooler or samples do not appear to have been compromised or tampered with.	True		2
Samples were received on ice.	True		3
Cooler Temperature is acceptable (</=6C, not frozen).	True		4
Cooler Temperature is recorded.	True		5
WV: Container Temperature is acceptable (</=6C, not frozen).	N/A		6
WV: Container Temperature is recorded.	N/A		7
COC is present.	True		8
COC is filled out in ink and legible.	True		9
COC is filled out with all pertinent information.	True		10
There are no discrepancies between the containers received and the COC.	True		11
Sample containers have legible labels.	True		12
Containers are not broken or leaking.	True		13
Sample collection date/times are provided.	True		14
Appropriate sample containers are used.	True		15
Sample bottles are completely filled.	True		16
There is sufficient vol. for all requested analyses.	True		
Is the Field Sampler's name present on COC?	True		
Sample custody seals are intact.	True		
VOA sample vials do not have headspace >6mm in diameter (none, if from WV)?	N/A		