

ANALYTICAL REPORT

PREPARED FOR

Attn: Jonathan Dippert
CT Male Associates DPC
50 Century Hill Dr
Latham, New York 12110

Generated 2/18/2025 12:13:40 PM

JOB DESCRIPTION

HFWTP 14.4756
HOO

JOB NUMBER

410-207155-1

Eurofins Lancaster Laboratories Environment Testing, LLC

Job Notes

This report may not be reproduced except in full, and with written approval from the laboratory. The results relate only to the samples tested. For questions please contact the Project Manager at the e-mail address or telephone number listed on this page.

Analytical test results meet all requirements of the associated regulatory program (i.e., NELAC (TNI), DoD, and ISO 17025) unless otherwise noted under the individual analysis.

Authorization



Generated
2/18/2025 12:13:40 PM

Authorized for release by
Kelly Gallagher, Project Manager
kelly.gallagher@et.eurofinsus.com
(717)205-7820

Compliance Statement

Analytical test results meet all requirements of the associated regulatory program (e.g., NELAC (TNI), DoD, and ISO 17025) unless otherwise noted under the individual analysis. Data qualifiers are applied to note exceptions. Noncompliant quality control (QC) is further explained in narrative comments.

- QC results that exceed the upper limits and are associated with non-detect samples are qualified but further narration is not required since the bias is high and does not change a non-detect result. Further narration is also not required with QC blank detection when the associated sample concentration is non-detect or more than ten times the level in the blank.
 - Matrix QC may not be reported if insufficient sample or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD is performed, unless otherwise specified in the method.
 - Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.
- Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

Measurement uncertainty values, as applicable, are available upon request.

Test results relate only to the sample tested. Clients should be aware that a critical step in a chemical or microbiological analysis is the collection of the sample. Unless the sample analyzed is truly representative of the bulk of material involved, the test results will be meaningless. If you have questions regarding the proper techniques of collecting samples, please contact us. We cannot be held responsible for sample integrity, however, unless sampling has been performed by a member of our staff. Times are local to the area of activity. Parameters listed in the 40 CFR Part 136 Table II as "analyze immediately" and tested in the laboratory are not performed within 15 minutes of collection.

This report shall not be reproduced except in full, without the written approval of the laboratory.

WARRANTY AND LIMITS OF LIABILITY - In accepting analytical work, we warrant the accuracy of test results for the sample as submitted. The foregoing express warranty is exclusive and is given in lieu of all other warranties, expressed or implied, except as otherwise agreed. We disclaim any other warranties, expressed or implied, including a warranty of fitness for particular purpose and warranty of merchantability. In no event shall Eurofins Lancaster Laboratories Environmental, LLC be liable for indirect, special, consequential, or incidental damages including, but not limited to, damages for loss of profit or goodwill regardless of (A) the negligence (either sole or concurrent) of Eurofins Lancaster Laboratories Environmental and (B) whether Eurofins Lancaster Laboratories Environmental has been informed of the possibility of such damages. We accept no legal responsibility for the purposes for which the client uses the test results. Except as otherwise agreed, no purchase order or other order for work shall be accepted by Eurofins Lancaster Laboratories Environmental which includes any conditions that vary from the Standard Terms and Conditions, and Eurofins Lancaster Laboratories Environmental hereby objects to any conflicting terms contained in any acceptance or order submitted by client.



Table of Contents

Cover Page 1

Table of Contents 4

Definitions/Glossary 5

Case Narrative 6

Detection Summary 7

Client Sample Results 8

Surrogate Summary 14

Isotope Dilution Summary 15

QC Sample Results 16

QC Association Summary 21

Lab Chronicle 23

Certification Summary 25

Method Summary 26

Sample Summary 27

Chain of Custody 28

Receipt Checklists 29



Definitions/Glossary

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Qualifiers

LCMS

Qualifier	Qualifier Description
cn	Refer to Case Narrative for further detail
J	Result is less than the RL but greater than or equal to the MDL and the concentration is an approximate value.
U	Indicates the analyte was analyzed for but not detected.

Glossary

Abbreviation	These commonly used abbreviations may or may not be present in this report.
☼	Listed under the "D" column to designate that the result is reported on a dry weight basis
%R	Percent Recovery
CFL	Contains Free Liquid
CFU	Colony Forming Unit
CNF	Contains No Free Liquid
DER	Duplicate Error Ratio (normalized absolute difference)
Dil Fac	Dilution Factor
DL	Detection Limit (DoD/DOE)
DL, RA, RE, IN	Indicates a Dilution, Re-analysis, Re-extraction, or additional Initial metals/anion analysis of the sample
DLC	Decision Level Concentration (Radiochemistry)
EDL	Estimated Detection Limit (Dioxin)
LOD	Limit of Detection (DoD/DOE)
LOQ	Limit of Quantitation (DoD/DOE)
MCL	EPA recommended "Maximum Contaminant Level"
MDA	Minimum Detectable Activity (Radiochemistry)
MDC	Minimum Detectable Concentration (Radiochemistry)
MDL	Method Detection Limit
ML	Minimum Level (Dioxin)
MPN	Most Probable Number
MQL	Method Quantitation Limit
NC	Not Calculated
ND	Not Detected at the reporting limit (or MDL or EDL if shown)
NEG	Negative / Absent
POS	Positive / Present
PQL	Practical Quantitation Limit
PRES	Presumptive
QC	Quality Control
RER	Relative Error Ratio (Radiochemistry)
RL	Reporting Limit or Requested Limit (Radiochemistry)
RPD	Relative Percent Difference, a measure of the relative difference between two points
TEF	Toxicity Equivalent Factor (Dioxin)
TEQ	Toxicity Equivalent Quotient (Dioxin)
TNTC	Too Numerous To Count

Case Narrative

Client: CT Male Associates DPC
Project: HFWTP 14.4756

Job ID: 410-207155-1

Job ID: 410-207155-1

Eurofins Lancaster Laboratories Environment

Job Narrative 410-207155-1

Analytical test results meet all requirements of the associated regulatory program listed on the Accreditation/Certification Summary Page unless otherwise noted under the individual analysis. Data qualifiers and/or narrative comments are included to explain any exceptions, if applicable.

- Matrix QC may not be reported if insufficient sample is provided or site-specific QC samples were not submitted. In these situations, to demonstrate precision and accuracy at a batch level, a LCS/LCSD may be performed, unless otherwise specified in the method.
- Surrogate and/or isotope dilution analyte recoveries (if applicable) which are outside of the QC window are confirmed unless attributed to a dilution or otherwise noted in the narrative.

Regulated compliance samples (e.g. SDWA, NPDES) must comply with the associated agency requirements/permits.

Receipt

The samples were received on 2/8/2025 9:40 AM. Unless otherwise noted below, the samples arrived in good condition, and, where required, properly preserved and on ice. The temperature of the cooler at receipt time was 2.5°C.

PFAS

Method 537.1_DW: The following sample was found to contain residual chlorine: GAC INFLUENT (410-207155-1).

No additional analytical or quality issues were noted, other than those described above or in the Definitions/ Glossary page.

Detection Summary

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Client Sample ID: GAC INFLUENT

Lab Sample ID: 410-207155-1

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorobutanoic acid	3.7		1.8	ng/L	1		537 (Mod)	Total/NA
Perfluorooctanesulfonamide	2.2		1.8	ng/L	1		537 (Mod)	Total/NA
Perfluoropentanoic acid	2.0		1.8	ng/L	1		537 (Mod)	Total/NA
Perfluoroheptanoic acid	7.3	cn	1.9	ng/L	1		EPA 537.1	Total/NA
Perfluorohexanoic acid	6.9	cn	1.9	ng/L	1		EPA 537.1	Total/NA
Perfluorooctanesulfonic acid	3.5	cn	1.9	ng/L	1		EPA 537.1	Total/NA
Perfluorooctanoic acid - DL	290	cn	19	ng/L	10		EPA 537.1	Total/NA

Client Sample ID: GAC MIDFLUENT

Lab Sample ID: 410-207155-2

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorobutanoic acid	1.8		1.7	ng/L	1		537 (Mod)	Total/NA

Client Sample ID: GAC EFFLUENT

Lab Sample ID: 410-207155-3

No Detections.

Client Sample ID: PV-1_75

Lab Sample ID: 410-207155-4

Analyte	Result	Qualifier	RL	Unit	Dil Fac	D	Method	Prep Type
Perfluorobutanoic acid	5.7		1.8	ng/L	1		537 (Mod)	Total/NA

Client Sample ID: SG1-FTB01-250207

Lab Sample ID: 410-207155-5

No Detections.

Client Sample ID: SG1-LTB01-250207

Lab Sample ID: 410-207155-6

No Detections.

This Detection Summary does not include radiochemical test results.

Eurofins Lancaster Laboratories Environment Testing, LLC

Client Sample Results

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Client Sample ID: GAC INFLUENT

Lab Sample ID: 410-207155-1

Date Collected: 02/07/25 10:05

Matrix: Water

Date Received: 02/08/25 09:40

Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:02	1
8:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:02	1
Perfluorobutanoic acid	3.7		1.8	ng/L		02/11/25 09:48	02/12/25 21:02	1
Perfluorodecanesulfonic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:02	1
Perfluoroheptanesulfonic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:02	1
Perfluorooctanesulfonamide	2.2		1.8	ng/L		02/11/25 09:48	02/12/25 21:02	1
Perfluoropentanoic acid	2.0		1.8	ng/L		02/11/25 09:48	02/12/25 21:02	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
M2-6:2 FTS	89		26 - 200	02/11/25 09:48	02/12/25 21:02	1
M2-8:2 FTS	76		27 - 200	02/11/25 09:48	02/12/25 21:02	1
13C4 PFBA	66		10 - 168	02/11/25 09:48	02/12/25 21:02	1
13C5 PFPeA	72		15 - 189	02/11/25 09:48	02/12/25 21:02	1
13C8 PFOS	81		44 - 153	02/11/25 09:48	02/12/25 21:02	1
13C8 FOSA	65		11 - 149	02/11/25 09:48	02/12/25 21:02	1
13C3 PFHxS	80		39 - 164	02/11/25 09:48	02/12/25 21:02	1

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	1.9	U cn	1.9	ng/L		02/11/25 07:03	02/14/25 06:57	1
NMeFOSAA	1.9	U cn	1.9	ng/L		02/11/25 07:03	02/14/25 06:57	1
Perfluorobutanesulfonic acid	1.9	U cn	1.9	ng/L		02/11/25 07:03	02/14/25 06:57	1
Perfluorodecanoic acid	1.9	U cn	1.9	ng/L		02/11/25 07:03	02/14/25 06:57	1
Perfluorododecanoic acid	1.9	U cn	1.9	ng/L		02/11/25 07:03	02/14/25 06:57	1
Perfluoroheptanoic acid	7.3	cn	1.9	ng/L		02/11/25 07:03	02/14/25 06:57	1
Perfluorohexanesulfonic acid	1.9	U cn	1.9	ng/L		02/11/25 07:03	02/14/25 06:57	1
Perfluorohexanoic acid	6.9	cn	1.9	ng/L		02/11/25 07:03	02/14/25 06:57	1
Perfluorononanoic acid	1.9	U cn	1.9	ng/L		02/11/25 07:03	02/14/25 06:57	1
Perfluorooctanesulfonic acid	3.5	cn	1.9	ng/L		02/11/25 07:03	02/14/25 06:57	1
Perfluorotetradecanoic acid	1.9	U cn	1.9	ng/L		02/11/25 07:03	02/14/25 06:57	1
Perfluorotridecanoic acid	1.9	U cn	1.9	ng/L		02/11/25 07:03	02/14/25 06:57	1
Perfluoroundecanoic acid	1.9	U cn	1.9	ng/L		02/11/25 07:03	02/14/25 06:57	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFDA	105	cn	70 - 130	02/11/25 07:03	02/14/25 06:57	1
13C2 PFHxA	103	cn	70 - 130	02/11/25 07:03	02/14/25 06:57	1
d5-NEtFOSAA	91	cn	70 - 130	02/11/25 07:03	02/14/25 06:57	1

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018 - DL

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
Perfluorooctanoic acid	290	cn	19	ng/L		02/11/25 07:03	02/17/25 15:01	10

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFDA	94	cn	70 - 130	02/11/25 07:03	02/17/25 15:01	10
13C2 PFHxA	92	cn	70 - 130	02/11/25 07:03	02/17/25 15:01	10
d5-NEtFOSAA	84	cn	70 - 130	02/11/25 07:03	02/17/25 15:01	10

Client Sample Results

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Client Sample ID: GAC MIDFLUENT

Lab Sample ID: 410-207155-2

Date Collected: 02/07/25 10:20

Matrix: Water

Date Received: 02/08/25 09:40

Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L		02/11/25 09:48	02/12/25 21:16	1
8:2 Fluorotelomer sulfonic acid	1.7	U	1.7	ng/L		02/11/25 09:48	02/12/25 21:16	1
Perfluorobutanoic acid	1.8		1.7	ng/L		02/11/25 09:48	02/12/25 21:16	1
Perfluorodecanesulfonic acid	1.7	U	1.7	ng/L		02/11/25 09:48	02/12/25 21:16	1
Perfluoroheptanesulfonic acid	1.7	U	1.7	ng/L		02/11/25 09:48	02/12/25 21:16	1
Perfluorooctanesulfonamide	1.7	U	1.7	ng/L		02/11/25 09:48	02/12/25 21:16	1
Perfluoropentanoic acid	1.7	U	1.7	ng/L		02/11/25 09:48	02/12/25 21:16	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
M2-6:2 FTS	81		26 - 200	02/11/25 09:48	02/12/25 21:16	1
M2-8:2 FTS	75		27 - 200	02/11/25 09:48	02/12/25 21:16	1
13C4 PFBA	59		10 - 168	02/11/25 09:48	02/12/25 21:16	1
13C5 PFPeA	56		15 - 189	02/11/25 09:48	02/12/25 21:16	1
13C8 PFOS	76		44 - 153	02/11/25 09:48	02/12/25 21:16	1
13C8 FOSA	63		11 - 149	02/11/25 09:48	02/12/25 21:16	1
13C3 PFHxS	73		39 - 164	02/11/25 09:48	02/12/25 21:16	1

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:02	1
NMeFOSAA	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:02	1
Perfluorobutanesulfonic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:02	1
Perfluorodecanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:02	1
Perfluorododecanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:02	1
Perfluoroheptanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:02	1
Perfluorohexanesulfonic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:02	1
Perfluorohexanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:02	1
Perfluorononanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:02	1
Perfluorooctanesulfonic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:02	1
Perfluorooctanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:02	1
Perfluorotetradecanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:02	1
Perfluorotridecanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:02	1
Perfluoroundecanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:02	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFDA	92		70 - 130	02/11/25 07:15	02/13/25 18:02	1
13C2 PFHxA	112		70 - 130	02/11/25 07:15	02/13/25 18:02	1
d5-NEtFOSAA	93		70 - 130	02/11/25 07:15	02/13/25 18:02	1

Client Sample Results

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Client Sample ID: GAC EFFLUENT

Lab Sample ID: 410-207155-3

Date Collected: 02/07/25 10:30

Matrix: Water

Date Received: 02/08/25 09:40

Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:29	1
8:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:29	1
Perfluorobutanoic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:29	1
Perfluorodecanesulfonic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:29	1
Perfluoroheptanesulfonic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:29	1
Perfluorooctanesulfonamide	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:29	1
Perfluoropentanoic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:29	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
M2-6:2 FTS	86		26 - 200	02/11/25 09:48	02/12/25 21:29	1
M2-8:2 FTS	75		27 - 200	02/11/25 09:48	02/12/25 21:29	1
13C4 PFBA	62		10 - 168	02/11/25 09:48	02/12/25 21:29	1
13C5 PFPeA	61		15 - 189	02/11/25 09:48	02/12/25 21:29	1
13C8 PFOS	85		44 - 153	02/11/25 09:48	02/12/25 21:29	1
13C8 FOSA	66		11 - 149	02/11/25 09:48	02/12/25 21:29	1
13C3 PFHxS	79		39 - 164	02/11/25 09:48	02/12/25 21:29	1

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:15	1
NMeFOSAA	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:15	1
Perfluorobutanesulfonic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:15	1
Perfluorodecanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:15	1
Perfluorododecanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:15	1
Perfluoroheptanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:15	1
Perfluorohexanesulfonic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:15	1
Perfluorohexanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:15	1
Perfluorononanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:15	1
Perfluorooctanesulfonic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:15	1
Perfluorooctanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:15	1
Perfluorotetradecanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:15	1
Perfluorotridecanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:15	1
Perfluoroundecanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:15	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFDA	90		70 - 130	02/11/25 07:15	02/13/25 18:15	1
13C2 PFHxA	106		70 - 130	02/11/25 07:15	02/13/25 18:15	1
d5-NEtFOSAA	92		70 - 130	02/11/25 07:15	02/13/25 18:15	1

Client Sample Results

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Client Sample ID: PV-1_75

Lab Sample ID: 410-207155-4

Date Collected: 02/07/25 10:40

Matrix: Water

Date Received: 02/08/25 09:40

Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:43	1
8:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:43	1
Perfluorobutanoic acid	5.7		1.8	ng/L		02/11/25 09:48	02/12/25 21:43	1
Perfluorodecanesulfonic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:43	1
Perfluoroheptanesulfonic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:43	1
Perfluorooctanesulfonamide	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:43	1
Perfluoropentanoic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:43	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
M2-6:2 FTS	91		26 - 200	02/11/25 09:48	02/12/25 21:43	1
M2-8:2 FTS	78		27 - 200	02/11/25 09:48	02/12/25 21:43	1
13C4 PFBA	67		10 - 168	02/11/25 09:48	02/12/25 21:43	1
13C5 PFPeA	62		15 - 189	02/11/25 09:48	02/12/25 21:43	1
13C8 PFOS	81		44 - 153	02/11/25 09:48	02/12/25 21:43	1
13C8 FOSA	65		11 - 149	02/11/25 09:48	02/12/25 21:43	1
13C3 PFHxS	79		39 - 164	02/11/25 09:48	02/12/25 21:43	1

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:42	1
NMeFOSAA	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:42	1
Perfluorobutanesulfonic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:42	1
Perfluorodecanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:42	1
Perfluorododecanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:42	1
Perfluoroheptanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:42	1
Perfluorohexanesulfonic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:42	1
Perfluorohexanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:42	1
Perfluorononanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:42	1
Perfluorooctanesulfonic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:42	1
Perfluorooctanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:42	1
Perfluorotetradecanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:42	1
Perfluorotridecanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:42	1
Perfluoroundecanoic acid	1.8	U	1.8	ng/L		02/11/25 07:15	02/13/25 18:42	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFDA	94		70 - 130	02/11/25 07:15	02/13/25 18:42	1
13C2 PFHxA	104		70 - 130	02/11/25 07:15	02/13/25 18:42	1
d5-NEtFOSAA	94		70 - 130	02/11/25 07:15	02/13/25 18:42	1

Client Sample Results

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Client Sample ID: SG1-FTB01-250207

Lab Sample ID: 410-207155-5

Date Collected: 02/07/25 10:50

Matrix: Water

Date Received: 02/08/25 09:40

Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:56	1
8:2 Fluorotelomer sulfonic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:56	1
Perfluorobutanoic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:56	1
Perfluorodecanesulfonic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:56	1
Perfluoroheptanesulfonic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:56	1
Perfluorooctanesulfonamide	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:56	1
Perfluoropentanoic acid	1.8	U	1.8	ng/L		02/11/25 09:48	02/12/25 21:56	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
M2-6:2 FTS	85		26 - 200	02/11/25 09:48	02/12/25 21:56	1
M2-8:2 FTS	74		27 - 200	02/11/25 09:48	02/12/25 21:56	1
13C4 PFBA	77		10 - 168	02/11/25 09:48	02/12/25 21:56	1
13C5 PFPeA	74		15 - 189	02/11/25 09:48	02/12/25 21:56	1
13C8 PFOS	84		44 - 153	02/11/25 09:48	02/12/25 21:56	1
13C8 FOSA	70		11 - 149	02/11/25 09:48	02/12/25 21:56	1
13C3 PFHxS	75		39 - 164	02/11/25 09:48	02/12/25 21:56	1

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	1.7	U	1.7	ng/L		02/11/25 07:15	02/13/25 19:09	1
NMeFOSAA	1.7	U	1.7	ng/L		02/11/25 07:15	02/13/25 19:09	1
Perfluorobutanesulfonic acid	1.7	U	1.7	ng/L		02/11/25 07:15	02/13/25 19:09	1
Perfluorodecanoic acid	1.7	U	1.7	ng/L		02/11/25 07:15	02/13/25 19:09	1
Perfluorododecanoic acid	1.7	U	1.7	ng/L		02/11/25 07:15	02/13/25 19:09	1
Perfluoroheptanoic acid	1.7	U	1.7	ng/L		02/11/25 07:15	02/13/25 19:09	1
Perfluorohexanesulfonic acid	1.7	U	1.7	ng/L		02/11/25 07:15	02/13/25 19:09	1
Perfluorohexanoic acid	1.7	U	1.7	ng/L		02/11/25 07:15	02/13/25 19:09	1
Perfluorononanoic acid	1.7	U	1.7	ng/L		02/11/25 07:15	02/13/25 19:09	1
Perfluorooctanesulfonic acid	1.7	U	1.7	ng/L		02/11/25 07:15	02/13/25 19:09	1
Perfluorooctanoic acid	1.7	U	1.7	ng/L		02/11/25 07:15	02/13/25 19:09	1
Perfluorotetradecanoic acid	1.7	U	1.7	ng/L		02/11/25 07:15	02/13/25 19:09	1
Perfluorotridecanoic acid	1.7	U	1.7	ng/L		02/11/25 07:15	02/13/25 19:09	1
Perfluoroundecanoic acid	1.7	U	1.7	ng/L		02/11/25 07:15	02/13/25 19:09	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFDA	94		70 - 130	02/11/25 07:15	02/13/25 19:09	1
13C2 PFHxA	101		70 - 130	02/11/25 07:15	02/13/25 19:09	1
d5-NEtFOSAA	99		70 - 130	02/11/25 07:15	02/13/25 19:09	1

Client Sample Results

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Client Sample ID: SG1-LTB01-250207

Lab Sample ID: 410-207155-6

Date Collected: 02/07/25 00:00

Matrix: Water

Date Received: 02/08/25 09:40

Method: EPA 537 (Mod) - EPA 537 Version 1.1 modified

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	1.9	U	1.9	ng/L		02/11/25 09:48	02/12/25 22:10	1
8:2 Fluorotelomer sulfonic acid	1.9	U	1.9	ng/L		02/11/25 09:48	02/12/25 22:10	1
Perfluorobutanoic acid	1.9	U	1.9	ng/L		02/11/25 09:48	02/12/25 22:10	1
Perfluorodecanesulfonic acid	1.9	U	1.9	ng/L		02/11/25 09:48	02/12/25 22:10	1
Perfluoroheptanesulfonic acid	1.9	U	1.9	ng/L		02/11/25 09:48	02/12/25 22:10	1
Perfluorooctanesulfonamide	1.9	U	1.9	ng/L		02/11/25 09:48	02/12/25 22:10	1
Perfluoropentanoic acid	1.9	U	1.9	ng/L		02/11/25 09:48	02/12/25 22:10	1

Isotope Dilution	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
M2-6:2 FTS	87		26 - 200	02/11/25 09:48	02/12/25 22:10	1
M2-8:2 FTS	71		27 - 200	02/11/25 09:48	02/12/25 22:10	1
13C4 PFBA	74		10 - 168	02/11/25 09:48	02/12/25 22:10	1
13C5 PFPeA	71		15 - 189	02/11/25 09:48	02/12/25 22:10	1
13C8 PFOS	79		44 - 153	02/11/25 09:48	02/12/25 22:10	1
13C8 FOSA	70		11 - 149	02/11/25 09:48	02/12/25 22:10	1
13C3 PFHxS	76		39 - 164	02/11/25 09:48	02/12/25 22:10	1

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Analyte	Result	Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 19:23	1
NMeFOSAA	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 19:23	1
Perfluorobutanesulfonic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 19:23	1
Perfluorodecanoic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 19:23	1
Perfluorododecanoic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 19:23	1
Perfluoroheptanoic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 19:23	1
Perfluorohexanesulfonic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 19:23	1
Perfluorohexanoic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 19:23	1
Perfluorononanoic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 19:23	1
Perfluorooctanesulfonic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 19:23	1
Perfluorooctanoic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 19:23	1
Perfluorotetradecanoic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 19:23	1
Perfluorotridecanoic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 19:23	1
Perfluoroundecanoic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 19:23	1

Surrogate	%Recovery	Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFDA	86		70 - 130	02/11/25 07:15	02/13/25 19:23	1
13C2 PFHxA	93		70 - 130	02/11/25 07:15	02/13/25 19:23	1
d5-NEtFOSAA	93		70 - 130	02/11/25 07:15	02/13/25 19:23	1

Surrogate Summary

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Surrogate Recovery (Acceptance Limits)		
		PFDA (70-130)	PFHxA (70-130)	d5NEFOS (70-130)
410-207155-1	GAC INFLUENT	105 cn	103 cn	91 cn
410-207155-1 - DL	GAC INFLUENT	94 cn	92 cn	84 cn
410-207155-2	GAC MIDFLUENT	92	112	93
410-207155-3	GAC EFFLUENT	90	106	92
410-207155-4	PV-1_75	94	104	94
410-207155-5	SG1-FTB01-250207	94	101	99
410-207155-6	SG1-LTB01-250207	86	93	93
LCS 410-604564/2-A	Lab Control Sample	96	101	92
LCS 410-604582/2-A	Lab Control Sample	87	108	83
LCSD 410-604564/3-A	Lab Control Sample Dup	97	103	90
LCSD 410-604582/3-A	Lab Control Sample Dup	94	106	93
LLCS 410-604564/4-A	Lab Control Sample	99	102	94
MB 410-604564/1-A	Method Blank	102	109	94
MB 410-604582/1-A	Method Blank	97	107	85

Surrogate Legend

PFDA = 13C2 PFDA

PFHxA = 13C2 PFHxA

d5NEFOS = d5-NEtFOSAA

Isotope Dilution Summary

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Method: 537 (Mod) - EPA 537 Version 1.1 modified

Matrix: Water

Prep Type: Total/NA

Lab Sample ID	Client Sample ID	Percent Isotope Dilution Recovery (Acceptance Limits)						
		M262FTS (26-200)	M282FTS (27-200)	PFBA (10-168)	PFPeA (15-189)	C8PFOS (44-153)	PFOSA (11-149)	C3PFHS (39-164)
410-207155-1	GAC INFLUENT	89	76	66	72	81	65	80
410-207155-2	GAC MIDFLUENT	81	75	59	56	76	63	73
410-207155-3	GAC EFFLUENT	86	75	62	61	85	66	79
410-207155-4	PV-1_75	91	78	67	62	81	65	79
410-207155-5	SG1-FTB01-250207	85	74	77	74	84	70	75
410-207155-6	SG1-LTB01-250207	87	71	74	71	79	70	76
LCS 410-604710/2-A	Lab Control Sample	92	80	75	75	87	73	80
LCSD 410-604710/3-A	Lab Control Sample Dup	86	69	69	70	81	69	74
MB 410-604710/1-A	Method Blank	83	72	68	72	80	66	74

Surrogate Legend

M262FTS = M2-6:2 FTS

M282FTS = M2-8:2 FTS

PFBA = 13C4 PFBA

PFPeA = 13C5 PFPeA

C8PFOS = 13C8 PFOS

PFOSA = 13C8 FOSA

C3PFHS = 13C3 PFHxS

QC Sample Results

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Method: 537 (Mod) - EPA 537 Version 1.1 modified

Lab Sample ID: MB 410-604710/1-A

Matrix: Water

Analysis Batch: 605195

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 604710

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
6:2 Fluorotelomer sulfonic acid	2.0	U	2.0	ng/L		02/11/25 09:48	02/12/25 18:33	1
8:2 Fluorotelomer sulfonic acid	2.0	U	2.0	ng/L		02/11/25 09:48	02/12/25 18:33	1
Perfluorobutanoic acid	2.0	U	2.0	ng/L		02/11/25 09:48	02/12/25 18:33	1
Perfluorodecanesulfonic acid	2.0	U	2.0	ng/L		02/11/25 09:48	02/12/25 18:33	1
Perfluoroheptanesulfonic acid	2.0	U	2.0	ng/L		02/11/25 09:48	02/12/25 18:33	1
Perfluorooctanesulfonamide	2.0	U	2.0	ng/L		02/11/25 09:48	02/12/25 18:33	1
Perfluoropentanoic acid	2.0	U	2.0	ng/L		02/11/25 09:48	02/12/25 18:33	1

Isotope Dilution	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
M2-6:2 FTS	83		26 - 200	02/11/25 09:48	02/12/25 18:33	1
M2-8:2 FTS	72		27 - 200	02/11/25 09:48	02/12/25 18:33	1
13C4 PFBA	68		10 - 168	02/11/25 09:48	02/12/25 18:33	1
13C5 PFPeA	72		15 - 189	02/11/25 09:48	02/12/25 18:33	1
13C8 PFOS	80		44 - 153	02/11/25 09:48	02/12/25 18:33	1
13C8 FOSA	66		11 - 149	02/11/25 09:48	02/12/25 18:33	1
13C3 PFHxS	74		39 - 164	02/11/25 09:48	02/12/25 18:33	1

Lab Sample ID: LCS 410-604710/2-A

Matrix: Water

Analysis Batch: 605195

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 604710

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
6:2 Fluorotelomer sulfonic acid	24.4	20.4		ng/L		84	56 - 130
8:2 Fluorotelomer sulfonic acid	24.6	23.6		ng/L		96	55 - 130
Perfluorobutanoic acid	25.6	30.8		ng/L		120	56 - 130
Perfluorodecanesulfonic acid	24.7	21.4		ng/L		87	53 - 130
Perfluoroheptanesulfonic acid	24.4	21.6		ng/L		89	55 - 130
Perfluorooctanesulfonamide	25.6	24.0		ng/L		94	64 - 133
Perfluoropentanoic acid	25.6	20.8		ng/L		81	56 - 130

Isotope Dilution	LCS %Recovery	LCS Qualifier	Limits
M2-6:2 FTS	92		26 - 200
M2-8:2 FTS	80		27 - 200
13C4 PFBA	75		10 - 168
13C5 PFPeA	75		15 - 189
13C8 PFOS	87		44 - 153
13C8 FOSA	73		11 - 149
13C3 PFHxS	80		39 - 164

Lab Sample ID: LCSD 410-604710/3-A

Matrix: Water

Analysis Batch: 605195

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 604710

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
6:2 Fluorotelomer sulfonic acid	24.4	19.6		ng/L		80	56 - 130	4	30
8:2 Fluorotelomer sulfonic acid	24.6	26.1		ng/L		106	55 - 130	10	30
Perfluorobutanoic acid	25.6	28.2		ng/L		110	56 - 130	9	30
Perfluorodecanesulfonic acid	24.7	22.2		ng/L		90	53 - 130	4	30
Perfluoroheptanesulfonic acid	24.4	21.9		ng/L		90	55 - 130	1	30

Eurofins Lancaster Laboratories Environment Testing, LLC

QC Sample Results

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Method: 537 (Mod) - EPA 537 Version 1.1 modified (Continued)

Lab Sample ID: LCSD 410-604710/3-A

Matrix: Water

Analysis Batch: 605195

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 604710

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
Perfluorooctanesulfonamide	25.6	24.5		ng/L		96	64 - 133	2	30
Perfluoropentanoic acid	25.6	20.8		ng/L		81	56 - 130	0	30

Isotope Dilution	LCSD %Recovery	LCSD Qualifier	Limits
M2-6:2 FTS	86		26 - 200
M2-8:2 FTS	69		27 - 200
13C4 PFBA	69		10 - 168
13C5 PFPeA	70		15 - 189
13C8 PFOS	81		44 - 153
13C8 FOSA	69		11 - 149
13C3 PFHxS	74		39 - 164

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018

Lab Sample ID: MB 410-604564/1-A

Matrix: Water

Analysis Batch: 605922

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 604564

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	2.0	U	2.0	ng/L		02/11/25 07:03	02/14/25 06:03	1
NMeFOSAA	2.0	U	2.0	ng/L		02/11/25 07:03	02/14/25 06:03	1
Perfluorobutanesulfonic acid	2.0	U	2.0	ng/L		02/11/25 07:03	02/14/25 06:03	1
Perfluorodecanoic acid	2.0	U	2.0	ng/L		02/11/25 07:03	02/14/25 06:03	1
Perfluorododecanoic acid	2.0	U	2.0	ng/L		02/11/25 07:03	02/14/25 06:03	1
Perfluoroheptanoic acid	2.0	U	2.0	ng/L		02/11/25 07:03	02/14/25 06:03	1
Perfluorohexanesulfonic acid	2.0	U	2.0	ng/L		02/11/25 07:03	02/14/25 06:03	1
Perfluorohexanoic acid	2.0	U	2.0	ng/L		02/11/25 07:03	02/14/25 06:03	1
Perfluorononanoic acid	2.0	U	2.0	ng/L		02/11/25 07:03	02/14/25 06:03	1
Perfluorooctanesulfonic acid	2.0	U	2.0	ng/L		02/11/25 07:03	02/14/25 06:03	1
Perfluorooctanoic acid	2.0	U	2.0	ng/L		02/11/25 07:03	02/14/25 06:03	1
Perfluorotetradecanoic acid	2.0	U	2.0	ng/L		02/11/25 07:03	02/14/25 06:03	1
Perfluorotridecanoic acid	2.0	U	2.0	ng/L		02/11/25 07:03	02/14/25 06:03	1
Perfluoroundecanoic acid	2.0	U	2.0	ng/L		02/11/25 07:03	02/14/25 06:03	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFDA	102		70 - 130	02/11/25 07:03	02/14/25 06:03	1
13C2 PFHxA	109		70 - 130	02/11/25 07:03	02/14/25 06:03	1
d5-NEtFOSAA	94		70 - 130	02/11/25 07:03	02/14/25 06:03	1

Lab Sample ID: LCS 410-604564/2-A

Matrix: Water

Analysis Batch: 605922

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 604564

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
NEtFOSAA	60.0	55.3		ng/L		92	70 - 130
NMeFOSAA	60.0	54.2		ng/L		90	70 - 130
Perfluorobutanesulfonic acid	53.1	52.6		ng/L		99	70 - 130
Perfluorodecanoic acid	60.0	57.7		ng/L		96	70 - 130
Perfluorododecanoic acid	60.0	53.2		ng/L		89	70 - 130

Eurofins Lancaster Laboratories Environment Testing, LLC

QC Sample Results

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018 (Continued)

Lab Sample ID: LCS 410-604564/2-A

Matrix: Water

Analysis Batch: 605922

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 604564

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluoroheptanoic acid	60.0	60.2		ng/L		100	70 - 130
Perfluorohexanesulfonic acid	54.8	54.8		ng/L		100	70 - 130
Perfluorohexanoic acid	60.0	58.6		ng/L		98	70 - 130
Perfluorononanoic acid	60.0	60.7		ng/L		101	70 - 130
Perfluorooctanesulfonic acid	55.6	53.3		ng/L		96	70 - 130
Perfluorooctanoic acid	60.0	59.1		ng/L		98	70 - 130
Perfluorotetradecanoic acid	60.0	52.7		ng/L		88	70 - 130
Perfluorotridecanoic acid	60.0	51.2		ng/L		85	70 - 130
Perfluoroundecanoic acid	60.0	57.2		ng/L		95	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
13C2 PFDA	96		70 - 130
13C2 PFHxA	101		70 - 130
d5-NEtFOSAA	92		70 - 130

Lab Sample ID: LCSD 410-604564/3-A

Matrix: Water

Analysis Batch: 605922

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 604564

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	Limit
NEtFOSAA	60.0	54.8		ng/L		91	70 - 130	1	30
NMeFOSAA	60.0	51.9		ng/L		87	70 - 130	4	30
Perfluorobutanesulfonic acid	53.1	53.1		ng/L		100	70 - 130	1	30
Perfluorodecanoic acid	60.0	59.6		ng/L		99	70 - 130	3	30
Perfluorododecanoic acid	60.0	52.0		ng/L		87	70 - 130	2	30
Perfluoroheptanoic acid	60.0	60.5		ng/L		101	70 - 130	0	30
Perfluorohexanesulfonic acid	54.8	54.4		ng/L		99	70 - 130	1	30
Perfluorohexanoic acid	60.0	63.2		ng/L		105	70 - 130	8	30
Perfluorononanoic acid	60.0	60.9		ng/L		101	70 - 130	0	30
Perfluorooctanesulfonic acid	55.6	53.2		ng/L		96	70 - 130	0	30
Perfluorooctanoic acid	60.0	60.2		ng/L		100	70 - 130	2	30
Perfluorotetradecanoic acid	60.0	52.4		ng/L		87	70 - 130	1	30
Perfluorotridecanoic acid	60.0	52.7		ng/L		88	70 - 130	3	30
Perfluoroundecanoic acid	60.0	57.8		ng/L		96	70 - 130	1	30

Surrogate	LCSD %Recovery	LCSD Qualifier	Limits
13C2 PFDA	97		70 - 130
13C2 PFHxA	103		70 - 130
d5-NEtFOSAA	90		70 - 130

Lab Sample ID: LLCS 410-604564/4-A

Matrix: Water

Analysis Batch: 605922

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 604564

Analyte	Spike Added	LLCS Result	LLCS Qualifier	Unit	D	%Rec	%Rec Limits
NEtFOSAA	1.92	1.91	J	ng/L		99	50 - 150
NMeFOSAA	1.92	1.80	J	ng/L		94	50 - 150
Perfluorobutanesulfonic acid	1.70	1.65	J	ng/L		97	50 - 150

Eurofins Lancaster Laboratories Environment Testing, LLC

QC Sample Results

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018 (Continued)

Lab Sample ID: LLCS 410-604564/4-A

Matrix: Water

Analysis Batch: 605922

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 604564

Analyte	Spike Added	LLCS Result	LLCS Qualifier	Unit	D	%Rec	%Rec Limits
Perfluorodecanoic acid	1.92	2.06		ng/L		107	50 - 150
Perfluorododecanoic acid	1.92	1.89	J	ng/L		98	50 - 150
Perfluoroheptanoic acid	1.92	2.03		ng/L		106	50 - 150
Perfluorohexanesulfonic acid	1.75	1.76	J	ng/L		101	50 - 150
Perfluorohexanoic acid	1.92	2.09		ng/L		109	50 - 150
Perfluorononanoic acid	1.92	2.10		ng/L		109	50 - 150
Perfluorooctanesulfonic acid	1.78	1.83	J	ng/L		103	50 - 150
Perfluorooctanoic acid	1.92	2.20		ng/L		114	50 - 150
Perfluorotetradecanoic acid	1.92	1.81	J	ng/L		94	50 - 150
Perfluorotridecanoic acid	1.92	1.81	J	ng/L		94	50 - 150
Perfluoroundecanoic acid	1.92	2.11		ng/L		110	50 - 150

Surrogate	LLCS %Recovery	LLCS Qualifier	Limits
13C2 PFDA	99		70 - 130
13C2 PFHxA	102		70 - 130
d5-NEtFOSAA	94		70 - 130

Lab Sample ID: MB 410-604582/1-A

Matrix: Water

Analysis Batch: 605683

Client Sample ID: Method Blank

Prep Type: Total/NA

Prep Batch: 604582

Analyte	MB Result	MB Qualifier	RL	Unit	D	Prepared	Analyzed	Dil Fac
NEtFOSAA	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 15:47	1
NMeFOSAA	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 15:47	1
Perfluorobutanesulfonic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 15:47	1
Perfluorodecanoic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 15:47	1
Perfluorododecanoic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 15:47	1
Perfluoroheptanoic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 15:47	1
Perfluorohexanesulfonic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 15:47	1
Perfluorohexanoic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 15:47	1
Perfluorononanoic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 15:47	1
Perfluorooctanesulfonic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 15:47	1
Perfluorooctanoic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 15:47	1
Perfluorotetradecanoic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 15:47	1
Perfluorotridecanoic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 15:47	1
Perfluoroundecanoic acid	2.0	U	2.0	ng/L		02/11/25 07:15	02/13/25 15:47	1

Surrogate	MB %Recovery	MB Qualifier	Limits	Prepared	Analyzed	Dil Fac
13C2 PFDA	97		70 - 130	02/11/25 07:15	02/13/25 15:47	1
13C2 PFHxA	107		70 - 130	02/11/25 07:15	02/13/25 15:47	1
d5-NEtFOSAA	85		70 - 130	02/11/25 07:15	02/13/25 15:47	1

Lab Sample ID: LCS 410-604582/2-A

Matrix: Water

Analysis Batch: 605683

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 604582

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
NEtFOSAA	20.5	17.0		ng/L		83	70 - 130

Eurofins Lancaster Laboratories Environment Testing, LLC

QC Sample Results

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Method: EPA 537.1 - EPA 537.1, Ver 1.0 Nov 2018 (Continued)

Lab Sample ID: LCS 410-604582/2-A

Matrix: Water

Analysis Batch: 605683

Client Sample ID: Lab Control Sample

Prep Type: Total/NA

Prep Batch: 604582

Analyte	Spike Added	LCS Result	LCS Qualifier	Unit	D	%Rec	%Rec Limits
NMeFOSAA	20.5	16.8		ng/L		82	70 - 130
Perfluorobutanesulfonic acid	18.1	18.4		ng/L		101	70 - 130
Perfluorodecanoic acid	20.5	17.4		ng/L		85	70 - 130
Perfluorododecanoic acid	20.5	15.5		ng/L		76	70 - 130
Perfluoroheptanoic acid	20.5	20.1		ng/L		98	70 - 130
Perfluorohexanesulfonic acid	18.7	18.1		ng/L		97	70 - 130
Perfluorohexanoic acid	20.5	20.7		ng/L		101	70 - 130
Perfluorononanoic acid	20.5	18.0		ng/L		88	70 - 130
Perfluorooctanesulfonic acid	19.0	17.6		ng/L		93	70 - 130
Perfluorooctanoic acid	20.5	18.8		ng/L		92	70 - 130
Perfluorotetradecanoic acid	20.5	14.8		ng/L		72	70 - 130
Perfluorotridecanoic acid	20.5	14.8		ng/L		72	70 - 130
Perfluoroundecanoic acid	20.5	17.6		ng/L		86	70 - 130

Surrogate	LCS %Recovery	LCS Qualifier	Limits
13C2 PFDA	87		70 - 130
13C2 PFHxA	108		70 - 130
d5-NEtFOSAA	83		70 - 130

Lab Sample ID: LCSD 410-604582/3-A

Matrix: Water

Analysis Batch: 605683

Client Sample ID: Lab Control Sample Dup

Prep Type: Total/NA

Prep Batch: 604582

Analyte	Spike Added	LCSD Result	LCSD Qualifier	Unit	D	%Rec	%Rec Limits	RPD	RPD Limit
NEtFOSAA	20.5	16.3		ng/L		80	70 - 130	5	30
NMeFOSAA	20.5	17.2		ng/L		84	70 - 130	2	30
Perfluorobutanesulfonic acid	18.1	15.9		ng/L		88	70 - 130	14	30
Perfluorodecanoic acid	20.5	19.0		ng/L		93	70 - 130	9	30
Perfluorododecanoic acid	20.5	17.3		ng/L		85	70 - 130	11	30
Perfluoroheptanoic acid	20.5	20.5		ng/L		100	70 - 130	2	30
Perfluorohexanesulfonic acid	18.7	17.8		ng/L		95	70 - 130	2	30
Perfluorohexanoic acid	20.5	20.0		ng/L		98	70 - 130	4	30
Perfluorononanoic acid	20.5	18.7		ng/L		91	70 - 130	4	30
Perfluorooctanesulfonic acid	19.0	17.7		ng/L		93	70 - 130	0	30
Perfluorooctanoic acid	20.5	19.2		ng/L		94	70 - 130	2	30
Perfluorotetradecanoic acid	20.5	16.4		ng/L		80	70 - 130	10	30
Perfluorotridecanoic acid	20.5	16.6		ng/L		81	70 - 130	11	30
Perfluoroundecanoic acid	20.5	18.3		ng/L		90	70 - 130	4	30

Surrogate	LCSD %Recovery	LCSD Qualifier	Limits
13C2 PFDA	94		70 - 130
13C2 PFHxA	106		70 - 130
d5-NEtFOSAA	93		70 - 130

QC Association Summary

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

LCMS

Prep Batch: 604564

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-207155-1	GAC INFLUENT	Total/NA	Water	537.1 DW Prep	
410-207155-1 - DL	GAC INFLUENT	Total/NA	Water	537.1 DW Prep	
MB 410-604564/1-A	Method Blank	Total/NA	Water	537.1 DW Prep	
LCS 410-604564/2-A	Lab Control Sample	Total/NA	Water	537.1 DW Prep	
LCSD 410-604564/3-A	Lab Control Sample Dup	Total/NA	Water	537.1 DW Prep	
LLCS 410-604564/4-A	Lab Control Sample	Total/NA	Water	537.1 DW Prep	

Prep Batch: 604582

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-207155-2	GAC MIDFLUENT	Total/NA	Water	537.1 DW Prep	
410-207155-3	GAC EFFLUENT	Total/NA	Water	537.1 DW Prep	
410-207155-4	PV-1_75	Total/NA	Water	537.1 DW Prep	
410-207155-5	SG1-FTB01-250207	Total/NA	Water	537.1 DW Prep	
410-207155-6	SG1-LTB01-250207	Total/NA	Water	537.1 DW Prep	
MB 410-604582/1-A	Method Blank	Total/NA	Water	537.1 DW Prep	
LCS 410-604582/2-A	Lab Control Sample	Total/NA	Water	537.1 DW Prep	
LCSD 410-604582/3-A	Lab Control Sample Dup	Total/NA	Water	537.1 DW Prep	

Prep Batch: 604710

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-207155-1	GAC INFLUENT	Total/NA	Water	SPE	
410-207155-2	GAC MIDFLUENT	Total/NA	Water	SPE	
410-207155-3	GAC EFFLUENT	Total/NA	Water	SPE	
410-207155-4	PV-1_75	Total/NA	Water	SPE	
410-207155-5	SG1-FTB01-250207	Total/NA	Water	SPE	
410-207155-6	SG1-LTB01-250207	Total/NA	Water	SPE	
MB 410-604710/1-A	Method Blank	Total/NA	Water	SPE	
LCS 410-604710/2-A	Lab Control Sample	Total/NA	Water	SPE	
LCSD 410-604710/3-A	Lab Control Sample Dup	Total/NA	Water	SPE	

Analysis Batch: 605195

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-207155-1	GAC INFLUENT	Total/NA	Water	537 (Mod)	604710
410-207155-2	GAC MIDFLUENT	Total/NA	Water	537 (Mod)	604710
410-207155-3	GAC EFFLUENT	Total/NA	Water	537 (Mod)	604710
410-207155-4	PV-1_75	Total/NA	Water	537 (Mod)	604710
410-207155-5	SG1-FTB01-250207	Total/NA	Water	537 (Mod)	604710
410-207155-6	SG1-LTB01-250207	Total/NA	Water	537 (Mod)	604710
MB 410-604710/1-A	Method Blank	Total/NA	Water	537 (Mod)	604710
LCS 410-604710/2-A	Lab Control Sample	Total/NA	Water	537 (Mod)	604710
LCSD 410-604710/3-A	Lab Control Sample Dup	Total/NA	Water	537 (Mod)	604710

Analysis Batch: 605683

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-207155-2	GAC MIDFLUENT	Total/NA	Water	EPA 537.1	604582
410-207155-3	GAC EFFLUENT	Total/NA	Water	EPA 537.1	604582
410-207155-4	PV-1_75	Total/NA	Water	EPA 537.1	604582
410-207155-5	SG1-FTB01-250207	Total/NA	Water	EPA 537.1	604582
410-207155-6	SG1-LTB01-250207	Total/NA	Water	EPA 537.1	604582
MB 410-604582/1-A	Method Blank	Total/NA	Water	EPA 537.1	604582
LCS 410-604582/2-A	Lab Control Sample	Total/NA	Water	EPA 537.1	604582

Eurofins Lancaster Laboratories Environment Testing, LLC

QC Association Summary

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

LCMS (Continued)

Analysis Batch: 605683 (Continued)

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
LCSD 410-604582/3-A	Lab Control Sample Dup	Total/NA	Water	EPA 537.1	604582

Analysis Batch: 605922

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-207155-1	GAC INFLUENT	Total/NA	Water	EPA 537.1	604564
MB 410-604564/1-A	Method Blank	Total/NA	Water	EPA 537.1	604564
LCS 410-604564/2-A	Lab Control Sample	Total/NA	Water	EPA 537.1	604564
LCSD 410-604564/3-A	Lab Control Sample Dup	Total/NA	Water	EPA 537.1	604564
LLCS 410-604564/4-A	Lab Control Sample	Total/NA	Water	EPA 537.1	604564

Analysis Batch: 606711

Lab Sample ID	Client Sample ID	Prep Type	Matrix	Method	Prep Batch
410-207155-1 - DL	GAC INFLUENT	Total/NA	Water	EPA 537.1	604564

Lab Chronicle

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Client Sample ID: GAC INFLUENT

Lab Sample ID: 410-207155-1

Date Collected: 02/07/25 10:05

Matrix: Water

Date Received: 02/08/25 09:40

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			604710	D5VP	ELLE	02/11/25 09:48
Total/NA	Analysis	537 (Mod)		1	605195	E9LB	ELLE	02/12/25 21:02
Total/NA	Prep	537.1 DW Prep			604564	DX7G	ELLE	02/11/25 07:03
Total/NA	Analysis	EPA 537.1		1	605922	MC7V	ELLE	02/14/25 06:57
Total/NA	Prep	537.1 DW Prep	DL		604564	DX7G	ELLE	02/11/25 07:03
Total/NA	Analysis	EPA 537.1	DL	10	606711	XBL5	ELLE	02/17/25 15:01

Client Sample ID: GAC MIDFLUENT

Lab Sample ID: 410-207155-2

Date Collected: 02/07/25 10:20

Matrix: Water

Date Received: 02/08/25 09:40

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			604710	D5VP	ELLE	02/11/25 09:48
Total/NA	Analysis	537 (Mod)		1	605195	E9LB	ELLE	02/12/25 21:16
Total/NA	Prep	537.1 DW Prep			604582	DX7G	ELLE	02/11/25 07:15
Total/NA	Analysis	EPA 537.1		1	605683	XBL5	ELLE	02/13/25 18:02

Client Sample ID: GAC EFFLUENT

Lab Sample ID: 410-207155-3

Date Collected: 02/07/25 10:30

Matrix: Water

Date Received: 02/08/25 09:40

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			604710	D5VP	ELLE	02/11/25 09:48
Total/NA	Analysis	537 (Mod)		1	605195	E9LB	ELLE	02/12/25 21:29
Total/NA	Prep	537.1 DW Prep			604582	DX7G	ELLE	02/11/25 07:15
Total/NA	Analysis	EPA 537.1		1	605683	XBL5	ELLE	02/13/25 18:15

Client Sample ID: PV-1_75

Lab Sample ID: 410-207155-4

Date Collected: 02/07/25 10:40

Matrix: Water

Date Received: 02/08/25 09:40

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			604710	D5VP	ELLE	02/11/25 09:48
Total/NA	Analysis	537 (Mod)		1	605195	E9LB	ELLE	02/12/25 21:43
Total/NA	Prep	537.1 DW Prep			604582	DX7G	ELLE	02/11/25 07:15
Total/NA	Analysis	EPA 537.1		1	605683	XBL5	ELLE	02/13/25 18:42

Client Sample ID: SG1-FTB01-250207

Lab Sample ID: 410-207155-5

Date Collected: 02/07/25 10:50

Matrix: Water

Date Received: 02/08/25 09:40

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			604710	D5VP	ELLE	02/11/25 09:48
Total/NA	Analysis	537 (Mod)		1	605195	E9LB	ELLE	02/12/25 21:56
Total/NA	Prep	537.1 DW Prep			604582	DX7G	ELLE	02/11/25 07:15
Total/NA	Analysis	EPA 537.1		1	605683	XBL5	ELLE	02/13/25 19:09

Eurofins Lancaster Laboratories Environment Testing, LLC

Lab Chronicle

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Client Sample ID: SG1-LTB01-250207
Date Collected: 02/07/25 00:00
Date Received: 02/08/25 09:40

Lab Sample ID: 410-207155-6
Matrix: Water

Prep Type	Batch Type	Batch Method	Run	Dilution Factor	Batch Number	Analyst	Lab	Prepared or Analyzed
Total/NA	Prep	SPE			604710	D5VP	ELLE	02/11/25 09:48
Total/NA	Analysis	537 (Mod)		1	605195	E9LB	ELLE	02/12/25 22:10
Total/NA	Prep	537.1 DW Prep			604582	DX7G	ELLE	02/11/25 07:15
Total/NA	Analysis	EPA 537.1		1	605683	XBL5	ELLE	02/13/25 19:23

Laboratory References:
ELLE = Eurofins Lancaster Laboratories Environment Testing, LLC, 2425 New Holland Pike, Lancaster, PA 17601, TEL (717)656-2300

Accreditation/Certification Summary

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Laboratory: Eurofins Lancaster Laboratories Environment Testing, LLC

Unless otherwise noted, all analytes for this laboratory were covered under each accreditation/certification below.

Authority	Program	Identification Number	Expiration Date
New York	NELAP	10670	04-01-25

The following analytes are included in this report, but the laboratory is not certified by the governing authority. This list may include analytes for which the agency does not offer certification.

Analysis Method	Prep Method	Matrix	Analyte
537 (Mod)	SPE	Water	6:2 Fluorotelomer sulfonic acid
537 (Mod)	SPE	Water	8:2 Fluorotelomer sulfonic acid
537 (Mod)	SPE	Water	Perfluorobutanoic acid
537 (Mod)	SPE	Water	Perfluorodecanesulfonic acid
537 (Mod)	SPE	Water	Perfluoroheptanesulfonic acid
537 (Mod)	SPE	Water	Perfluorooctanesulfonamide
537 (Mod)	SPE	Water	Perfluoropentanoic acid

Method Summary

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Method	Method Description	Protocol	Laboratory
537 (Mod)	EPA 537 Version 1.1 modified	EPA	ELLE
EPA 537.1	EPA 537.1, Ver 1.0 Nov 2018	EPA	ELLE
537.1 DW Prep	Extraction of Perfluorinated Alkyl Acids	EPA	ELLE
SPE	PFAS by SPE	Lab SOP	ELLE

Protocol References:

EPA = US Environmental Protection Agency
Lab SOP = Laboratory Standard Operating Procedure

Laboratory References:

ELLE = Eurofins Lancaster Laboratories Environment Testing, LLC, 2425 New Holland Pike, Lancaster, PA 17601, TEL (717)656-2300

Sample Summary

Client: CT Male Associates DPC
Project/Site: HFWTP 14.4756

Job ID: 410-207155-1
SDG: HOO

Lab Sample ID	Client Sample ID	Matrix	Collected	Received
410-207155-1	GAC INFLUENT	Water	02/07/25 10:05	02/08/25 09:40
410-207155-2	GAC MIDFLUENT	Water	02/07/25 10:20	02/08/25 09:40
410-207155-3	GAC EFFLUENT	Water	02/07/25 10:30	02/08/25 09:40
410-207155-4	PV-1_75	Water	02/07/25 10:40	02/08/25 09:40
410-207155-5	SG1-FTB01-250207	Water	02/07/25 10:50	02/08/25 09:40
410-207155-6	SG1-LTB01-250207	Water	02/07/25 00:00	02/08/25 09:40



410-207155 Chain of Custody

Chain of Custody Record

818453729371



Sampler: Dan Achtyl		Lab PM: Kelly Gallagher		Carrier Tracking No(s):		COC No:	
Phone: 518-786-7400		E-Mail: Kelly.Gallagher@et.eurofinsus.com		State of Origin: NY		Page: 1 of 1	
Christopher Ormsby, Jonathan Lippert		PWSID:		Analysis Requested		Job #:	
Company: CT Male Associates DPC		Due Date Requested:		Field Filtered Sample (Yes or No) Perform MS/MSD (Yes or No) PFC_IDA - (MOD) 7 PFAS Compounds 537_DW - 14 PFAS Drinking Water List		Preservation Codes: A - HCL M - Hexane B - NaOH N - None C - Zn Acetate O - AsNaO2 D - Nitric Acid P - Na2O4S E - NaHSO4 Q - Na2SO3 F - MeOH R - Na2S2O3 G - Amchlor S - H2SO4 H - Ascorbic Acid T - TSP Dodecahydrate I - Ice U - Acetone J - DI Water V - MCAA K - EDTA W - pH 4-5 L - EDA Z - other (specify)	
Address: 50 Century Hill Dr		TAT Requested (days): Standard					
City: Latham		Compliance Project: ☐ Yes ☐ No					
State, Zip: New York, 12110		PO #: 14.4756					
Phone: 518-786-7400		WO #:					
Email: c.ormsby@ctmale.com, j.dippert@ctmale.com		Project #:		Total Number of containers		Other: Y - Trizma	
Project Name: HFWTP 14.4756		SSOW#:					
Site: HFWTP							
Sample Identification		Sample Date	Sample Time	Sample Type (C=comp, G=grab)	Matrix (W=water, S=solid, O=waste/oil, BT=Tissue, A=Air)	Special Instructions/Note:	
				Preservation Code:			
✓ GAC INFLUENT	2/7/25	1005	G	W	N N x x	8	PFAS Batch QC Sample Here
✓ GAC MIDFLUENT	2/7/25	1030	G	W	N N x x	4	
✓ GAC EFFLUENT	2/7/25	1030	G	W	N N x x	4	
✓ PV-1_75	2/7/25	1040	G	W	N N x x	4	
✓ SG1-FTB01-250207	2/7/25	1050	G	W	N N x x	4	
✓ SG1-LTB01-250207	2/7/25	—	G	W	N N x x	4	
Possible Hazard Identification		Sample Disposal (A fee may be assessed if samples are retained longer than 1 month)					
☐ Non-Hazard ☐ Flammable ☐ Skin Irritant ☐ Poison B ☒ Unknown ☐ Radiological		☐ Return To Client ☒ Disposal By Lab ☐ Archive For _____ Months					
Deliverable Requested: I, II, III, IV, Other (specify)		EQUIS 1-File; ASP-B		Special Instructions/QC Requirements:			
Empty Kit Relinquished by:		Date:		Time:		Method of Shipment:	
Relinquished by:		Date/Time: 2/7/2025 1230		Company: CTM		Received by:	
Relinquished by:		Date/Time:		Company:		Received by:	
Relinquished by:		Date/Time:		Company:		Received by:	
Custody Seals Intact: ☒ Yes ☐ No		Custody Seal No.:		Cooler Temperature(s) °C and Other Remarks: 10.2°C - 2.5			

Ver: 01/16/2019

Login Sample Receipt Checklist

Client: CT Male Associates DPC

Job Number: 410-207155-1

SDG Number: HOO

Login Number: 207155

List Source: Eurofins Lancaster Laboratories Environment Testing, LLC

List Number: 1

Creator: Ballard, Megan

Question	Answer	Comment
The cooler's custody seal is intact.	True	
The cooler or samples do not appear to have been compromised or tampered with.	True	
Samples were received on ice.	True	
Cooler Temperature acceptable,where thermal pres is required(</=6C, not frozen).	True	
Cooler Temperature is recorded.	True	
WV:Container Temp acceptable,where thermal pres is required (</=6C, not frozen).	N/A	
WV: Container Temperature is recorded.	N/A	
COC is present.	True	
COC is filled out in ink and legible.	True	
COC is filled out with all pertinent information.	True	
There are no discrepancies between the containers received and the COC.	True	
Sample containers have legible labels.	True	
Containers are not broken or leaking.	True	
Sample collection date/times are provided.	True	
Appropriate sample containers are used.	True	
Sample bottles are completely filled.	True	
There is sufficient vol. for all requested analyses.	True	
Is the Field Sampler's name present on COC?	True	
Sample custody seals are intact.	True	
VOA sample vials do not have headspace >6mm in diameter (none, if from WV)?	N/A	

Sample Preservation Checks (performed by the laboratory)

Question	Answer	Comment
Did the sample containers checked meet expected preservation conditions?	False	Refer to Job Narrative for details.